

October 17, 2016

Nancy Henner
Hannibal CSD
928 Cayuga St.
Hannibal, NY 13074

RE: Project: Han CSD - Fairly Elementary
Pace Project No.: 10365060

Dear Nancy Henner:

Enclosed are the analytical results for sample(s) received by the laboratory on October 06, 2016. The results relate only to the samples included in this report. Results reported herein conform to the most current, applicable TNI/NELAC standards and the laboratory's Quality Assurance Manual, where applicable, unless otherwise noted in the body of the report.

Samples, in the electronic data deliverable (EDD) that accompanied this report, were flagged yellow if they exceeded the NYSDOH 15 ppb action level.

The space designated for the sampler on the chain of custody (COC) indicates if these samples were collected by Pace Analytical or were collected by the school district and just relinquished to Pace Analytical.

If you have any questions concerning this report, please feel free to contact me.

Sincerely,



Amanda Albrecht
amanda.albrecht@pacelabs.com
Project Manager

Enclosures



REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

CERTIFICATIONS

Project: Han CSD - Fairly Elementary

Pace Project No.: 10365060

Minnesota Certification IDs

1700 Elm Street SE Suite 200, Minneapolis, MN 55414

Alaska Certification UST-107

525 N 8th Street, Salina, KS 67401

A2LA Certification #: 2926.01

Alaska Certification #: UST-078

Alaska Certification #MN00064

Alabama Certification #40770

Arizona Certification #: AZ-0014

Arkansas Certification #: 88-0680

California Certification #: 01155CA

Colorado Certification #Pace

Connecticut Certification #: PH-0256

EPA Region 8 Certification #: 8TMS-L

Florida/NELAP Certification #: E87605

Guam Certification #:14-008r

Georgia Certification #: 959

Georgia EPD #: Pace

Idaho Certification #: MN00064

Hawaii Certification #MN00064

Illinois Certification #: 200011

Indiana Certification#C-MN-01

Iowa Certification #: 368

Kansas Certification #: E-10167

Kentucky Dept of Envi. Protection - DW #90062

Kentucky Dept of Envi. Protection - WW #:90062

Louisiana DEQ Certification #: 3086

Louisiana DHH #: LA140001

Maine Certification #: 2013011

Maryland Certification #: 322

Michigan DEPH Certification #: 9909

Minnesota Certification #: 027-053-137

Mississippi Certification #: Pace

Montana Certification #: MT0092

Nevada Certification #: MN_00064

Nebraska Certification #: Pace

New Jersey Certification #: MN-002

New York Certification #: 11647

North Carolina Certification #: 530

North Carolina State Public Health #: 27700

North Dakota Certification #: R-036

Ohio EPA #: 4150

Ohio VAP Certification #: CL101

Oklahoma Certification #: 9507

Oregon Certification #: MN200001

Oregon Certification #: MN300001

Pennsylvania Certification #: 68-00563

Puerto Rico Certification

Saipan (CNMI) #:MP0003

South Carolina #:74003001

Texas Certification #: T104704192

Tennessee Certification #: 02818

Utah Certification #: MN000642013-4

Virginia DGS Certification #: 251

Virginia/VELAP Certification #: Pace

Washington Certification #: C486

West Virginia Certification #: 382

West Virginia DHHR #:9952C

Wisconsin Certification #: 999407970

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

SAMPLE SUMMARY

Project: Han CSD - Fairly Elementary

Pace Project No.: 10365060

Lab ID	Sample ID	Matrix	Date Collected	Date Received
10365060001	Kitchen/Disposal Sprayer	Water	09/27/16 02:31	10/06/16 10:15
10365060002	Cafeteria/Girls Room	Water	09/27/16 02:34	10/06/16 10:15
10365060003	Cafeteria/Boys Room	Water	09/27/16 02:34	10/06/16 10:15
10365060004	Room 040/ CS	Water	09/27/16 02:36	10/06/16 10:15
10365060005	Room 030/ CS	Water	09/27/16 02:37	10/06/16 10:15
10365060006	Room 010/ CS	Water	09/27/16 02:37	10/06/16 10:15
10365060007	Room 010/ BS	Water	09/27/16 02:38	10/06/16 10:15
10365060008	Room 020/ CS	Water	09/27/16 02:38	10/06/16 10:15
10365060009	Room 020/ BS	Water	09/27/16 02:40	10/06/16 10:15
10365060010	MAIN ENT./LADIES RM/LEFT SINK	Water	09/27/16 02:43	10/06/16 10:15
10365060011	MAIN ENT./LADIES RM/RIGHT SINK	Water	09/27/16 02:44	10/06/16 10:15
10365060012	MAIN ENT./MENS RM/LEFT SINK	Water	09/27/16 02:45	10/06/16 10:15
10365060013	MAIN ENT./MENS RM/RIGHT SINK	Water	09/27/16 02:45	10/06/16 10:15
10365060014	ROOM 107/CS	Water	09/27/16 02:46	10/06/16 10:15
10365060015	ROOM 105/CS	Water	09/27/16 02:46	10/06/16 10:15
10365060016	ROOM 106/CS	Water	09/27/16 02:47	10/06/16 10:15
10365060017	ROOM 103/CS	Water	09/27/16 02:47	10/06/16 10:15
10365060018	ROOM 104/CS	Water	09/27/16 02:48	10/06/16 10:15
10365060019	ROOM 102/CS	Water	09/27/16 02:50	10/06/16 10:15
10365060020	ROOM 102/BS	Water	09/27/16 02:50	10/06/16 10:15
10365060021	ROOM 101/CS	Water	09/27/16 02:51	10/06/16 10:15
10365060022	ROOM 101/BS	Water	09/27/16 02:52	10/06/16 10:15
10365060023	TEACHERS LOUNGE	Water	09/27/16 02:52	10/06/16 10:15
10365060024	RESPECT RD/MENS RM SINK	Water	09/27/16 02:56	10/06/16 10:15
10365060025	RESPECT RD/LADIES RM SINK	Water	09/27/16 02:59	10/06/16 10:15
10365060026	ROOM 126/CS	Water	09/27/16 02:59	10/06/16 10:15
10365060027	ROOM 121/CS	Water	09/27/16 03:00	10/06/16 10:15
10365060028	ROOM 123/CS	Water	09/27/16 03:01	10/06/16 10:15
10365060029	ROOM 122/CS	Water	09/27/16 03:02	10/06/16 10:15
10365060030	ROOM 125/CS	Water	09/27/16 03:03	10/06/16 10:15
10365060031	ROOM 124/CS	Water	09/27/16 03:04	10/06/16 10:15
10365060032	ROOM 126/CS	Water	09/27/16 03:04	10/06/16 10:15
10365060033	ROOM 127/CS	Water	09/27/16 03:05	10/06/16 10:15
10365060034	ROOM 130/CS	Water	09/27/16 03:06	10/06/16 10:15
10365060035	ROOM 130/BS	Water	09/27/16 03:07	10/06/16 10:15
10365060036	ROOM 131/BS-LEFT	Water	09/27/16 03:09	10/06/16 10:15
10365060037	ROOM 131/BS-RIGHT	Water	09/27/16 03:10	10/06/16 10:15

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

SAMPLE SUMMARY

Project: Han CSD - Fairly Elementary
Pace Project No.: 10365060

Lab ID	Sample ID	Matrix	Date Collected	Date Received
10365060038	BOYS LOCKER ROOM/LEFT SINK	Water	09/27/16 03:12	10/06/16 10:15
10365060039	BOYS LOCKER ROOM/RIGHT SINK	Water	09/27/16 03:12	10/06/16 10:15
10365060040	BOYS GYM/OFFICE BATH	Water	09/27/16 03:14	10/06/16 10:15
10365060041	ROOM 132/BS	Water	09/27/16 03:15	10/06/16 10:15
10365060042	MENS RM/ACROSS132/SINK1	Water	09/27/16 03:19	10/06/16 10:15
10365060043	MENS RM/ACROSS132/SINK2	Water	09/27/16 03:20	10/06/16 10:15
10365060044	MENS RM/ACROSS132/SINK3	Water	09/27/16 03:20	10/06/16 10:15
10365060045	MENS RM/ACROSS132/SINK4	Water	09/27/16 03:21	10/06/16 10:15
10365060046	LadiesRM/ACROSS132/SINK1	Water	09/27/16 03:26	10/06/16 10:15
10365060047	LadiesRM/ACROSS132/SINK2	Water	09/27/16 03:26	10/06/16 10:15
10365060048	LadiesRM/ACROSS132/SINK3	Water	09/27/16 03:27	10/06/16 10:15
10365060049	LadiesRM/ACROSS132/SINK4	Water	09/27/16 03:27	10/06/16 10:15
10365060050	ROOM 133/LEFT SINK	Water	09/27/16 03:30	10/06/16 10:15
10365060051	ROOM 133/RIGHT SINK	Water	09/27/16 03:30	10/06/16 10:15
10365060052	ROOM 133/BS	Water	09/27/16 03:32	10/06/16 10:15
10365060053	GIRLS LOCKER ROOM/LEFT SINK	Water	09/27/16 03:34	10/06/16 10:15
10365060054	GIRLS LOCKER ROOM/RIGHT SINK	Water	09/27/16 03:34	10/06/16 10:15
10365060055	GIRLS COACHES BATH	Water	09/27/16 03:35	10/06/16 10:15
10365060056	ROOM 134/CS	Water	09/27/16 03:36	10/06/16 10:15
10365060057	ROOM 134/BS	Water	09/27/16 03:36	10/06/16 10:15
10365060058	ROOM 136/CS	Water	09/27/16 03:41	10/06/16 10:15
10365060059	ROOM 136/BS	Water	09/27/16 03:42	10/06/16 10:15
10365060060	ROOM 135/BS	Water	09/27/16 03:42	10/06/16 10:15
10365060061	ROOM 135/CS	Water	09/27/16 03:41	10/06/16 10:15
10365060062	ROOM 137/CS	Water	09/27/16 03:43	10/06/16 10:15
10365060063	ROOM 137/BS	Water	09/27/16 03:44	10/06/16 10:15
10365060064	ROOM 138/CS	Water	09/27/16 03:45	10/06/16 10:15
10365060065	ROOM 138/BS	Water	09/27/16 03:45	10/06/16 10:15
10365060066	ROOM 140/CS	Water	09/27/16 03:46	10/06/16 10:15
10365060067	ROOM 140/BS	Water	09/27/16 03:46	10/06/16 10:15
10365060068	ROOM 139/CS	Water	09/27/16 03:47	10/06/16 10:15
10365060069	ROOM 139/BS	Water	09/27/16 03:47	10/06/16 10:15
10365060070	ROOM 141/CS	Water	09/27/16 03:50	10/06/16 10:15
10365060071	ROOM 141/BS	Water	09/27/16 03:50	10/06/16 10:15
10365060072	ROOM 201/CS	Water	09/27/16 03:55	10/06/16 10:15
10365060073	ROOM 202/CS	Water	09/27/16 03:55	10/06/16 10:15

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

SAMPLE SUMMARY

Project: Han CSD - Fairly Elementary

Pace Project No.: 10365060

Lab ID	Sample ID	Matrix	Date Collected	Date Received
10365060074	ROOM 203/CS	Water	09/27/16 03:56	10/06/16 10:15
10365060075	ROOM 204/CS	Water	09/27/16 03:56	10/06/16 10:15
10365060076	ROOM 205/CS	Water	09/27/16 03:57	10/06/16 10:15
10365060077	2NDFLR/BOYS ROOM/SINK1	Water	09/27/16 03:59	10/06/16 10:15
10365060078	2NDFLR/BOYS ROOM/SINK2	Water	09/27/16 04:00	10/06/16 10:15
10365060079	2NDFLR/BOYS ROOM/SINK3	Water	09/27/16 04:01	10/06/16 10:15
10365060080	2NDFLR/BOYS ROOM/SINK4	Water	09/27/16 04:02	10/06/16 10:15
10365060081	2NDFLR/BOYS ROOM/SINK5	Water	09/27/16 04:02	10/06/16 10:15
10365060082	2NDFLR/BOYS ROOM/SINK6	Water	09/27/16 04:03	10/06/16 10:15
10365060083	2NDFLR/FACULTY BATH/LEFT SINK	Water	09/27/16 04:06	10/06/16 10:15
10365060084	2NDFLR/FACULTY BATH/RIGHT SINK	Water	09/27/16 04:06	10/06/16 10:15
10365060085	ROOM 206/CS	Water	09/27/16 04:08	10/06/16 10:15
10365060086	ROOM 207/CS	Water	09/27/16 04:09	10/06/16 10:15
10365060087	2ND FLR/WOMENS FACULTY/LEFT	Water	09/27/16 04:10	10/06/16 10:15
10365060088	2ND FLR/WOMENS FACULTY/RIGHT	Water	09/27/16 04:11	10/06/16 10:15
10365060089	2ND FLR/GIRLS ROOM/SINK 1	Water	09/27/16 04:12	10/06/16 10:15
10365060090	2ND FLR/GIRLS ROOM/SINK 2	Water	09/27/16 04:13	10/06/16 10:15
10365060091	2ND FLR/GIRLS ROOM/SINK 3	Water	09/27/16 04:14	10/06/16 10:15
10365060092	2ND FLR/GIRLS ROOM/SINK 4	Water	09/27/16 04:15	10/06/16 10:15
10365060093	2ND FLR/GIRLS ROOM/SINK 5	Water	09/27/16 04:16	10/06/16 10:15
10365060094	2ND FLR/GIRLS ROOM/SINK 6	Water	09/27/16 04:17	10/06/16 10:15
10365060095	ROOM 209/CS	Water	09/27/16 04:17	10/06/16 10:15
10365060096	ROOM 208/CS	Water	09/27/16 04:18	10/06/16 10:15
10365060097	ROOM 210/CS	Water	09/27/16 04:20	10/06/16 10:15
10365060098	ROOM 212/CS	Water	09/27/16 04:21	10/06/16 10:15
10365060099	ROOM 211/CS	Water	09/27/16 04:21	10/06/16 10:15
10365060100	ROOM 213/CS	Water	09/27/16 04:22	10/06/16 10:15
10365060101	ROOM 215/CS	Water	09/27/16 04:24	10/06/16 10:15
10365060102	ROOM 214/CS	Water	09/27/16 04:25	10/06/16 10:15
10365060103	ROOM 216/CS	Water	09/27/16 04:25	10/06/16 10:15

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

SAMPLE ANALYTE COUNT

Project: Han CSD - Fairly Elementary

Pace Project No.: 10365060

Lab ID	Sample ID	Method	Analysts	Analytes Reported	Laboratory
10365060001	Kitchen/Disposal Sprayer	EPA 200.8	TT3	1	PASI-M
10365060002	Cafeteria/Girls Room	EPA 200.8	TT3	1	PASI-M
10365060003	Cafeteria/Boys Room	EPA 200.8	TT3	1	PASI-M
10365060004	Room 040/ CS	EPA 200.8	TT3	1	PASI-M
10365060005	Room 030/ CS	EPA 200.8	TT3	1	PASI-M
10365060006	Room 010/ CS	EPA 200.8	TT3	1	PASI-M
10365060007	Room 010/ BS	EPA 200.8	TT3	1	PASI-M
10365060008	Room 020/ CS	EPA 200.8	TT3	1	PASI-M
10365060009	Room 020/ BS	EPA 200.8	TT3	1	PASI-M
10365060010	MAIN ENT./LADIES RM/LEFT SINK	EPA 200.8	TT3	1	PASI-M
10365060011	MAIN ENT./LADIES RM/RIGHT SINK	EPA 200.8	TT3	1	PASI-M
10365060012	MAIN ENT./MENS RM/LEFT SINK	EPA 200.8	TT3	1	PASI-M
10365060013	MAIN ENT./MENS RM/RIGHT SINK	EPA 200.8	TT3	1	PASI-M
10365060014	ROOM 107/CS	EPA 200.8	TT3	1	PASI-M
10365060015	ROOM 105/CS	EPA 200.8	TT3	1	PASI-M
10365060016	ROOM 106/CS	EPA 200.8	TT3	1	PASI-M
10365060017	ROOM 103/CS	EPA 200.8	TT3	1	PASI-M
10365060018	ROOM 104/CS	EPA 200.8	TT3	1	PASI-M
10365060019	ROOM 102/CS	EPA 200.8	TT3	1	PASI-M
10365060020	ROOM 102/BS	EPA 200.8	TT3	1	PASI-M
10365060021	ROOM 101/CS	EPA 200.8	WBS	1	PASI-M
10365060022	ROOM 101/BS	EPA 200.8	WBS	1	PASI-M
10365060023	TEACHERS LOUNGE	EPA 200.8	WBS	1	PASI-M
10365060024	RESPECT RD/MENS RM SINK	EPA 200.8	WBS	1	PASI-M
10365060025	RESPECT RD/LADIES RM SINK	EPA 200.8	WBS	1	PASI-M
10365060026	ROOM 126/CS	EPA 200.8	WBS	1	PASI-M
10365060027	ROOM 121/CS	EPA 200.8	WBS	1	PASI-M
10365060028	ROOM 123/CS	EPA 200.8	TT3	1	PASI-M
10365060029	ROOM 122/CS	EPA 200.8	WBS	1	PASI-M
10365060030	ROOM 125/CS	EPA 200.8	WBS	1	PASI-M
10365060031	ROOM 124/CS	EPA 200.8	WBS	1	PASI-M
10365060032	ROOM 126/CS	EPA 200.8	WBS	1	PASI-M
10365060033	ROOM 127/CS	EPA 200.8	WBS	1	PASI-M
10365060034	ROOM 130/CS	EPA 200.8	TT3	1	PASI-M
10365060035	ROOM 130/BS	EPA 200.8	WBS	1	PASI-M
10365060036	ROOM 131/BS-LEFT	EPA 200.8	WBS	1	PASI-M
10365060037	ROOM 131/BS-RIGHT	EPA 200.8	WBS	1	PASI-M

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

SAMPLE ANALYTE COUNT

Project: Han CSD - Fairly Elementary

Pace Project No.: 10365060

Lab ID	Sample ID	Method	Analysts	Analytes Reported	Laboratory
10365060038	BOYS LOCKER ROOM/LEFT SINK	EPA 200.8	WBS	1	PASI-M
10365060039	BOYS LOCKER ROOM/RIGHT SINK	EPA 200.8	WBS	1	PASI-M
10365060040	BOYS GYM/OFFICE BATH	EPA 200.8	WBS	1	PASI-M
10365060041	ROOM 132/BS	EPA 200.8	WBS	1	PASI-M
10365060042	MENS RM/ACROSS132/SINK1	EPA 200.8	WBS	1	PASI-M
10365060043	MENS RM/ACROSS132/SINK2	EPA 200.8	WBS	1	PASI-M
10365060044	MENS RM/ACROSS132/SINK3	EPA 200.8	WBS	1	PASI-M
10365060045	MENS RM/ACROSS132/SINK4	EPA 200.8	WBS	1	PASI-M
10365060046	LadiesRM/ACROSS132/SINK1	EPA 200.8	WBS	1	PASI-M
10365060047	LadiesRM/ACROSS132/SINK2	EPA 200.8	WBS	1	PASI-M
10365060048	LadiesRM/ACROSS132/SINK3	EPA 200.8	WBS	1	PASI-M
10365060049	LadiesRM/ACROSS132/SINK4	EPA 200.8	WBS	1	PASI-M
10365060050	ROOM 133/LEFT SINK	EPA 200.8	WBS	1	PASI-M
10365060051	ROOM 133/RIGHT SINK	EPA 200.8	WBS	1	PASI-M
10365060052	ROOM 133/BS	EPA 200.8	WBS	1	PASI-M
10365060053	GIRLS LOCKER ROOM/LEFT SINK	EPA 200.8	WBS	1	PASI-M
10365060054	GIRLS LOCKER ROOM/RIGHT SINK	EPA 200.8	WBS	1	PASI-M
10365060055	GIRLS COACHES BATH	EPA 200.8	WBS	1	PASI-M
10365060056	ROOM 134/CS	EPA 200.8	WBS	1	PASI-M
10365060057	ROOM 134/BS	EPA 200.8	WBS	1	PASI-M
10365060058	ROOM 136/CS	EPA 200.8	WBS	1	PASI-M
10365060059	ROOM 136/BS	EPA 200.8	WBS	1	PASI-M
10365060060	ROOM 135/BS	EPA 200.8	WBS	1	PASI-M
10365060061	ROOM 135/CS	EPA 200.8	TT3	1	PASI-M
10365060062	ROOM 137/CS	EPA 200.8	TT3	1	PASI-M
10365060063	ROOM 137/BS	EPA 200.8	TT3	1	PASI-M
10365060064	ROOM 138/CS	EPA 200.8	TT3	1	PASI-M
10365060065	ROOM 138/BS	EPA 200.8	TT3	1	PASI-M
10365060066	ROOM 140/CS	EPA 200.8	TT3	1	PASI-M
10365060067	ROOM 140/BS	EPA 200.8	TT3	1	PASI-M
10365060068	ROOM 139/CS	EPA 200.8	TT3	1	PASI-M
10365060069	ROOM 139/BS	EPA 200.8	TT3	1	PASI-M
10365060070	ROOM 141/CS	EPA 200.8	TT3	1	PASI-M
10365060071	ROOM 141/BS	EPA 200.8	TT3	1	PASI-M
10365060072	ROOM 201/CS	EPA 200.8	TT3	1	PASI-M
10365060073	ROOM 202/CS	EPA 200.8	TT3	1	PASI-M
10365060074	ROOM 203/CS	EPA 200.8	TT3	1	PASI-M

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

SAMPLE ANALYTE COUNT

Project: Han CSD - Fairly Elementary
Pace Project No.: 10365060

Lab ID	Sample ID	Method	Analysts	Analytes Reported	Laboratory
10365060075	ROOM 204/CS	EPA 200.8	TT3	1	PASI-M
10365060076	ROOM 205/CS	EPA 200.8	TT3	1	PASI-M
10365060077	2NDFLR/BOYS ROOM/SINK1	EPA 200.8	TT3	1	PASI-M
10365060078	2NDFLR/BOYS ROOM/SINK2	EPA 200.8	TT3	1	PASI-M
10365060079	2NDFLR/BOYS ROOM/SINK3	EPA 200.8	TT3	1	PASI-M
10365060080	2NDFLR/BOYS ROOM/SINK4	EPA 200.8	TT3	1	PASI-M
10365060081	2NDFLR/BOYS ROOM/SINK5	EPA 200.8	WBS	1	PASI-M
10365060082	2NDFLR/BOYS ROOM/SINK6	EPA 200.8	WBS	1	PASI-M
10365060083	2NDFLR/FACULTY BATH/LEFT SINK	EPA 200.8	WBS	1	PASI-M
10365060084	2NDFLR/FACULTY BATH/RIGHT SINK	EPA 200.8	WBS	1	PASI-M
10365060085	ROOM 206/CS	EPA 200.8	WBS	1	PASI-M
10365060086	ROOM 207/CS	EPA 200.8	WBS	1	PASI-M
10365060087	2ND FLR/WOMENS FACULTY/LEFT	EPA 200.8	WBS	1	PASI-M
10365060088	2ND FLR/WOMENS FACULTY/RIGHT	EPA 200.8	WBS	1	PASI-M
10365060089	2ND FLR/GIRLS ROOM/SINK 1	EPA 200.8	WBS	1	PASI-M
10365060090	2ND FLR/GIRLS ROOM/SINK 2	EPA 200.8	WBS	1	PASI-M
10365060091	2ND FLR/GIRLS ROOM/SINK 3	EPA 200.8	WBS	1	PASI-M
10365060092	2ND FLR/GIRLS ROOM/SINK 4	EPA 200.8	WBS	1	PASI-M
10365060093	2ND FLR/GIRLS ROOM/SINK 5	EPA 200.8	TT3	1	PASI-M
10365060094	2ND FLR/GIRLS ROOM/SINK 6	EPA 200.8	WBS	1	PASI-M
10365060095	ROOM 209/CS	EPA 200.8	WBS	1	PASI-M
10365060096	ROOM 208/CS	EPA 200.8	WBS	1	PASI-M
10365060097	ROOM 210/CS	EPA 200.8	WBS	1	PASI-M
10365060098	ROOM 212/CS	EPA 200.8	WBS	1	PASI-M
10365060099	ROOM 211/CS	EPA 200.8	WBS	1	PASI-M
10365060100	ROOM 213/CS	EPA 200.8	WBS	1	PASI-M
10365060101	ROOM 215/CS	EPA 200.8	WBS	1	PASI-M
10365060102	ROOM 214/CS	EPA 200.8	WBS	1	PASI-M
10365060103	ROOM 216/CS	EPA 200.8	WBS	1	PASI-M

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

ANALYTICAL RESULTS

Project: Han CSD - Fairly Elementary
Pace Project No.: 10365060

Sample: Kitchen/Disposal Sprayer		Lab ID: 10365060001	Collected: 09/27/16 02:31	Received: 10/06/16 10:15	Matrix: Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS, DW		Analytical Method: EPA 200.8						
Lead	10.9	ug/L	0.10	1		10/11/16 17:28	7439-92-1	
Sample: Cafeteria/Girls Room		Lab ID: 10365060002	Collected: 09/27/16 02:34	Received: 10/06/16 10:15	Matrix: Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS, DW		Analytical Method: EPA 200.8						
Lead	0.64	ug/L	0.10	1		10/11/16 17:38	7439-92-1	
Sample: Cafeteria/Boys Room		Lab ID: 10365060003	Collected: 09/27/16 02:34	Received: 10/06/16 10:15	Matrix: Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS, DW		Analytical Method: EPA 200.8						
Lead	0.32	ug/L	0.10	1		10/11/16 17:40	7439-92-1	
Sample: Room 040/ CS		Lab ID: 10365060004	Collected: 09/27/16 02:36	Received: 10/06/16 10:15	Matrix: Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS, DW		Analytical Method: EPA 200.8						
Lead	0.25	ug/L	0.10	1		10/11/16 17:41	7439-92-1	
Sample: Room 030/ CS		Lab ID: 10365060005	Collected: 09/27/16 02:37	Received: 10/06/16 10:15	Matrix: Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS, DW		Analytical Method: EPA 200.8						
Lead	0.64	ug/L	0.10	1		10/11/16 17:43	7439-92-1	
Sample: Room 010/ CS		Lab ID: 10365060006	Collected: 09/27/16 02:37	Received: 10/06/16 10:15	Matrix: Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS, DW		Analytical Method: EPA 200.8						
Lead	0.70	ug/L	0.10	1		10/11/16 17:44	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

ANALYTICAL RESULTS

Project: Han CSD - Fairly Elementary

Pace Project No.: 10365060

Sample: Room 010/ BS		Lab ID: 10365060007	Collected: 09/27/16 02:38	Received: 10/06/16 10:15	Matrix: Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS, DW		Analytical Method: EPA 200.8						
Lead	0.78	ug/L	0.10	1		10/11/16 17:46	7439-92-1	
Sample: Room 020/ CS		Lab ID: 10365060008	Collected: 09/27/16 02:38	Received: 10/06/16 10:15	Matrix: Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS, DW		Analytical Method: EPA 200.8						
Lead	0.26	ug/L	0.10	1		10/11/16 17:47	7439-92-1	
Sample: Room 020/ BS		Lab ID: 10365060009	Collected: 09/27/16 02:40	Received: 10/06/16 10:15	Matrix: Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS, DW		Analytical Method: EPA 200.8						
Lead	0.65	ug/L	0.10	1		10/11/16 17:52	7439-92-1	
Sample: MAIN ENT./LADIES RM/LEFT SINK		Lab ID: 10365060010	Collected: 09/27/16 02:43	Received: 10/06/16 10:15	Matrix: Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS, DW		Analytical Method: EPA 200.8						
Lead	1.1	ug/L	0.10	1		10/11/16 17:53	7439-92-1	
Sample: MAIN ENT./LADIES RM/RIGHT SINK		Lab ID: 10365060011	Collected: 09/27/16 02:44	Received: 10/06/16 10:15	Matrix: Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS, DW		Analytical Method: EPA 200.8						
Lead	0.37	ug/L	0.10	1		10/11/16 17:55	7439-92-1	
Sample: MAIN ENT./MENS RM/LEFT SINK		Lab ID: 10365060012	Collected: 09/27/16 02:45	Received: 10/06/16 10:15	Matrix: Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS, DW		Analytical Method: EPA 200.8						
Lead	0.79	ug/L	0.10	1		10/11/16 17:58	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

ANALYTICAL RESULTS

Project: Han CSD - Fairly Elementary

Pace Project No.: 10365060

Sample: MAIN ENT./MENS RM/RIGHT SINK		Lab ID: 10365060013	Collected: 09/27/16 02:45	Received: 10/06/16 10:15	Matrix: Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS, DW		Analytical Method: EPA 200.8						
Lead	1.9	ug/L	0.10	1		10/11/16 17:59	7439-92-1	
Sample: ROOM 107/CS		Lab ID: 10365060014	Collected: 09/27/16 02:46	Received: 10/06/16 10:15	Matrix: Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS, DW		Analytical Method: EPA 200.8						
Lead	0.50	ug/L	0.10	1		10/11/16 18:01	7439-92-1	
Sample: ROOM 105/CS		Lab ID: 10365060015	Collected: 09/27/16 02:46	Received: 10/06/16 10:15	Matrix: Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS, DW		Analytical Method: EPA 200.8						
Lead	0.49	ug/L	0.10	1		10/11/16 18:03	7439-92-1	
Sample: ROOM 106/CS		Lab ID: 10365060016	Collected: 09/27/16 02:47	Received: 10/06/16 10:15	Matrix: Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS, DW		Analytical Method: EPA 200.8						
Lead	0.22	ug/L	0.10	1		10/11/16 18:04	7439-92-1	
Sample: ROOM 103/CS		Lab ID: 10365060017	Collected: 09/27/16 02:47	Received: 10/06/16 10:15	Matrix: Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS, DW		Analytical Method: EPA 200.8						
Lead	0.84	ug/L	0.10	1		10/11/16 18:06	7439-92-1	
Sample: ROOM 104/CS		Lab ID: 10365060018	Collected: 09/27/16 02:48	Received: 10/06/16 10:15	Matrix: Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS, DW		Analytical Method: EPA 200.8						
Lead	0.32	ug/L	0.10	1		10/11/16 18:13	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

ANALYTICAL RESULTS

Project: Han CSD - Fairly Elementary

Pace Project No.: 10365060

Sample: ROOM 102/CS	Lab ID: 10365060019	Collected: 09/27/16 02:50	Received: 10/06/16 10:15	Matrix: Water				
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual

200.8 MET ICPMS, DW	Analytical Method: EPA 200.8							
Lead	0.30	ug/L	0.10	1		10/11/16 18:14	7439-92-1	

Sample: ROOM 102/BS	Lab ID: 10365060020	Collected: 09/27/16 02:50	Received: 10/06/16 10:15	Matrix: Water				
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual

200.8 MET ICPMS, DW	Analytical Method: EPA 200.8							
Lead	2.7	ug/L	0.10	1		10/11/16 18:16	7439-92-1	

Sample: ROOM 101/CS	Lab ID: 10365060021	Collected: 09/27/16 02:51	Received: 10/06/16 10:15	Matrix: Water				
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual

200.8 MET ICPMS, DW	Analytical Method: EPA 200.8							
Lead	0.39	ug/L	0.10	1		10/11/16 19:45	7439-92-1	

Sample: ROOM 101/BS	Lab ID: 10365060022	Collected: 09/27/16 02:52	Received: 10/06/16 10:15	Matrix: Water				
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual

200.8 MET ICPMS, DW	Analytical Method: EPA 200.8							
Lead	1.4	ug/L	0.10	1		10/11/16 19:54	7439-92-1	

Sample: TEACHERS LOUNGE	Lab ID: 10365060023	Collected: 09/27/16 02:52	Received: 10/06/16 10:15	Matrix: Water				
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual

200.8 MET ICPMS, DW	Analytical Method: EPA 200.8							
Lead	1.5	ug/L	0.10	1		10/11/16 19:56	7439-92-1	

Sample: RESPECT RD/MENS RM SINK	Lab ID: 10365060024	Collected: 09/27/16 02:56	Received: 10/06/16 10:15	Matrix: Water				
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual

200.8 MET ICPMS, DW	Analytical Method: EPA 200.8							
Lead	4.9	ug/L	0.10	1		10/11/16 19:58	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

ANALYTICAL RESULTS

Project: Han CSD - Fairly Elementary

Pace Project No.: 10365060

Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
Sample: RESPECT RD/LADIES RM SINK								
		Lab ID: 10365060025	Collected: 09/27/16 02:59	Received: 10/06/16 10:15	Matrix: Water			
Analytical Method: EPA 200.8								
Lead	3.9	ug/L	0.10	1		10/11/16 20:04	7439-92-1	
Sample: ROOM 126/CS								
		Lab ID: 10365060026	Collected: 09/27/16 02:59	Received: 10/06/16 10:15	Matrix: Water			
Analytical Method: EPA 200.8								
Lead	10.9	ug/L	0.10	1		10/11/16 20:05	7439-92-1	
Sample: ROOM 121/CS								
		Lab ID: 10365060027	Collected: 09/27/16 03:00	Received: 10/06/16 10:15	Matrix: Water			
Analytical Method: EPA 200.8								
Lead	5.6	ug/L	0.10	1		10/11/16 20:07	7439-92-1	
Sample: ROOM 123/CS								
		Lab ID: 10365060028	Collected: 09/27/16 03:01	Received: 10/06/16 10:15	Matrix: Water			
Analytical Method: EPA 200.8 Preparation Method: EPA 200.8								
Lead	7.4	ug/L	0.10	1	10/12/16 09:30	10/12/16 17:11	7439-92-1	
Sample: ROOM 122/CS								
		Lab ID: 10365060029	Collected: 09/27/16 03:02	Received: 10/06/16 10:15	Matrix: Water			
Analytical Method: EPA 200.8								
Lead	2.6	ug/L	0.10	1		10/11/16 20:10	7439-92-1	
Sample: ROOM 125/CS								
		Lab ID: 10365060030	Collected: 09/27/16 03:03	Received: 10/06/16 10:15	Matrix: Water			
Analytical Method: EPA 200.8								
Lead	4.9	ug/L	0.10	1		10/11/16 20:11	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

ANALYTICAL RESULTS

Project: Han CSD - Fairly Elementary
Pace Project No.: 10365060

Sample: ROOM 124/CS		Lab ID: 10365060031	Collected: 09/27/16 03:04	Received: 10/06/16 10:15	Matrix: Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS, DW		Analytical Method: EPA 200.8						
Lead	5.7	ug/L	0.10	1		10/11/16 20:12	7439-92-1	
Sample: ROOM 126/CS		Lab ID: 10365060032	Collected: 09/27/16 03:04	Received: 10/06/16 10:15	Matrix: Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS, DW		Analytical Method: EPA 200.8						
Lead	9.1	ug/L	0.10	1		10/11/16 20:14	7439-92-1	
Sample: ROOM 127/CS		Lab ID: 10365060033	Collected: 09/27/16 03:05	Received: 10/06/16 10:15	Matrix: Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS, DW		Analytical Method: EPA 200.8						
Lead	9.1	ug/L	0.10	1		10/11/16 20:16	7439-92-1	
Sample: ROOM 130/CS		Lab ID: 10365060034	Collected: 09/27/16 03:06	Received: 10/06/16 10:15	Matrix: Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS		Analytical Method: EPA 200.8 Preparation Method: EPA 200.8						
Lead	2.9	ug/L	0.10	1	10/12/16 09:30	10/12/16 17:12	7439-92-1	
Sample: ROOM 130/BS		Lab ID: 10365060035	Collected: 09/27/16 03:07	Received: 10/06/16 10:15	Matrix: Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS, DW		Analytical Method: EPA 200.8						
Lead	0.59	ug/L	0.10	1		10/11/16 20:18	7439-92-1	
Sample: ROOM 131/BS-LEFT		Lab ID: 10365060036	Collected: 09/27/16 03:09	Received: 10/06/16 10:15	Matrix: Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS, DW		Analytical Method: EPA 200.8						
Lead	1.1	ug/L	0.10	1		10/11/16 20:22	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

ANALYTICAL RESULTS

Project: Han CSD - Fairly Elementary

Pace Project No.: 10365060

Sample: ROOM 131/BS-RIGHT		Lab ID: 10365060037	Collected: 09/27/16 03:10	Received: 10/06/16 10:15	Matrix: Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS, DW		Analytical Method: EPA 200.8						
Lead	0.59	ug/L	0.10	1		10/11/16 20:23	7439-92-1	
Sample: BOYS LOCKER ROOM/LEFT SINK		Lab ID: 10365060038	Collected: 09/27/16 03:12	Received: 10/06/16 10:15	Matrix: Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS, DW		Analytical Method: EPA 200.8						
Lead	0.74	ug/L	0.10	1		10/11/16 20:24	7439-92-1	
Sample: BOYS LOCKER ROOM/RIGHT SINK		Lab ID: 10365060039	Collected: 09/27/16 03:12	Received: 10/06/16 10:15	Matrix: Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS, DW		Analytical Method: EPA 200.8						
Lead	0.77	ug/L	0.10	1		10/11/16 20:26	7439-92-1	
Sample: BOYS GYM/OFFICE BATH		Lab ID: 10365060040	Collected: 09/27/16 03:14	Received: 10/06/16 10:15	Matrix: Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS, DW		Analytical Method: EPA 200.8						
Lead	0.74	ug/L	0.10	1		10/11/16 20:27	7439-92-1	
Sample: ROOM 132/BS		Lab ID: 10365060041	Collected: 09/27/16 03:15	Received: 10/06/16 10:15	Matrix: Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS, DW		Analytical Method: EPA 200.8						
Lead	0.87	ug/L	0.10	1		10/11/16 19:01	7439-92-1	
Sample: MENS RM/ACROSS132/SINK1		Lab ID: 10365060042	Collected: 09/27/16 03:19	Received: 10/06/16 10:15	Matrix: Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS, DW		Analytical Method: EPA 200.8						
Lead	0.80	ug/L	0.10	1		10/11/16 19:10	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

ANALYTICAL RESULTS

Project: Han CSD - Fairly Elementary

Pace Project No.: 10365060

Sample: MENS RM/ACROSS132/SINK2		Lab ID: 10365060043	Collected: 09/27/16 03:20	Received: 10/06/16 10:15	Matrix: Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual

200.8 MET ICPMS, DW		Analytical Method: EPA 200.8						
Lead	0.82	ug/L	0.10	1		10/11/16 19:11	7439-92-1	

Sample: MENS RM/ACROSS132/SINK3		Lab ID: 10365060044	Collected: 09/27/16 03:20	Received: 10/06/16 10:15	Matrix: Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual

200.8 MET ICPMS, DW		Analytical Method: EPA 200.8						
Lead	0.84	ug/L	0.10	1		10/11/16 19:12	7439-92-1	

Sample: MENS RM/ACROSS132/SINK4		Lab ID: 10365060045	Collected: 09/27/16 03:21	Received: 10/06/16 10:15	Matrix: Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual

200.8 MET ICPMS, DW		Analytical Method: EPA 200.8						
Lead	0.49	ug/L	0.10	1		10/11/16 19:14	7439-92-1	

Sample: LadiesRM/ACROSS132/SINK1		Lab ID: 10365060046	Collected: 09/27/16 03:26	Received: 10/06/16 10:15	Matrix: Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual

200.8 MET ICPMS, DW		Analytical Method: EPA 200.8						
Lead	1.2	ug/L	0.10	1		10/11/16 19:15	7439-92-1	

Sample: LadiesRM/ACROSS132/SINK2		Lab ID: 10365060047	Collected: 09/27/16 03:26	Received: 10/06/16 10:15	Matrix: Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual

200.8 MET ICPMS, DW		Analytical Method: EPA 200.8						
Lead	0.90	ug/L	0.10	1		10/11/16 19:16	7439-92-1	

Sample: LadiesRM/ACROSS132/SINK3		Lab ID: 10365060048	Collected: 09/27/16 03:27	Received: 10/06/16 10:15	Matrix: Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual

200.8 MET ICPMS, DW		Analytical Method: EPA 200.8						
Lead	0.94	ug/L	0.10	1		10/11/16 19:18	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

ANALYTICAL RESULTS

Project: Han CSD - Fairly Elementary
Pace Project No.: 10365060

Sample: LadiesRM/ACROSS132/SIN K4		Lab ID: 10365060049	Collected: 09/27/16 03:27	Received: 10/06/16 10:15	Matrix: Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS, DW		Analytical Method: EPA 200.8						
Lead	0.58	ug/L	0.10	1		10/11/16 19:19	7439-92-1	
Sample: ROOM 133/LEFT SINK		Lab ID: 10365060050	Collected: 09/27/16 03:30	Received: 10/06/16 10:15	Matrix: Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS, DW		Analytical Method: EPA 200.8						
Lead	3.1	ug/L	0.10	1		10/11/16 19:20	7439-92-1	
Sample: ROOM 133/RIGHT SINK		Lab ID: 10365060051	Collected: 09/27/16 03:30	Received: 10/06/16 10:15	Matrix: Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS, DW		Analytical Method: EPA 200.8						
Lead	1.4	ug/L	0.10	1		10/11/16 19:25	7439-92-1	
Sample: ROOM 133/BS		Lab ID: 10365060052	Collected: 09/27/16 03:32	Received: 10/06/16 10:15	Matrix: Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS, DW		Analytical Method: EPA 200.8						
Lead	1.2	ug/L	0.10	1		10/11/16 19:27	7439-92-1	
Sample: GIRLS LOCKER ROOM/LEFT SINK		Lab ID: 10365060053	Collected: 09/27/16 03:34	Received: 10/06/16 10:15	Matrix: Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS, DW		Analytical Method: EPA 200.8						
Lead	0.95	ug/L	0.10	1		10/11/16 19:29	7439-92-1	
Sample: GIRLS LOCKER ROOM/RIGHT SINK		Lab ID: 10365060054	Collected: 09/27/16 03:34	Received: 10/06/16 10:15	Matrix: Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS, DW		Analytical Method: EPA 200.8						
Lead	0.80	ug/L	0.10	1		10/11/16 19:30	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

ANALYTICAL RESULTS

Project: Han CSD - Fairly Elementary
Pace Project No.: 10365060

Sample: GIRLS COACHES BATH		Lab ID: 10365060055	Collected: 09/27/16 03:35	Received: 10/06/16 10:15	Matrix: Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS, DW		Analytical Method: EPA 200.8						
Lead	0.70	ug/L	0.10	1		10/11/16 19:31	7439-92-1	
Sample: ROOM 134/CS		Lab ID: 10365060056	Collected: 09/27/16 03:36	Received: 10/06/16 10:15	Matrix: Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS, DW		Analytical Method: EPA 200.8						
Lead	1.2	ug/L	0.10	1		10/11/16 19:32	7439-92-1	
Sample: ROOM 134/BS		Lab ID: 10365060057	Collected: 09/27/16 03:36	Received: 10/06/16 10:15	Matrix: Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS, DW		Analytical Method: EPA 200.8						
Lead	1.5	ug/L	0.10	1		10/11/16 19:34	7439-92-1	
Sample: ROOM 136/CS		Lab ID: 10365060058	Collected: 09/27/16 03:41	Received: 10/06/16 10:15	Matrix: Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS, DW		Analytical Method: EPA 200.8						
Lead	0.64	ug/L	0.10	1		10/11/16 19:35	7439-92-1	
Sample: ROOM 136/BS		Lab ID: 10365060059	Collected: 09/27/16 03:42	Received: 10/06/16 10:15	Matrix: Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS, DW		Analytical Method: EPA 200.8						
Lead	0.56	ug/L	0.10	1		10/11/16 19:37	7439-92-1	
Sample: ROOM 135/BS		Lab ID: 10365060060	Collected: 09/27/16 03:42	Received: 10/06/16 10:15	Matrix: Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS, DW		Analytical Method: EPA 200.8						
Lead	2.0	ug/L	0.10	1		10/11/16 19:41	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

ANALYTICAL RESULTS

Project: Han CSD - Fairly Elementary
Pace Project No.: 10365060

Sample: ROOM 135/CS		Lab ID: 10365060061	Collected: 09/27/16 03:41	Received: 10/06/16 10:15	Matrix: Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS, DW		Analytical Method: EPA 200.8						
Lead	0.64	ug/L	0.10	1		10/10/16 14:13	7439-92-1	
Sample: ROOM 137/CS		Lab ID: 10365060062	Collected: 09/27/16 03:43	Received: 10/06/16 10:15	Matrix: Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS, DW		Analytical Method: EPA 200.8						
Lead	0.73	ug/L	0.10	1		10/10/16 14:19	7439-92-1	
Sample: ROOM 137/BS		Lab ID: 10365060063	Collected: 09/27/16 03:44	Received: 10/06/16 10:15	Matrix: Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS, DW		Analytical Method: EPA 200.8						
Lead	2.3	ug/L	0.10	1		10/10/16 14:21	7439-92-1	
Sample: ROOM 138/CS		Lab ID: 10365060064	Collected: 09/27/16 03:45	Received: 10/06/16 10:15	Matrix: Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS, DW		Analytical Method: EPA 200.8						
Lead	1.7	ug/L	0.10	1		10/10/16 14:22	7439-92-1	
Sample: ROOM 138/BS		Lab ID: 10365060065	Collected: 09/27/16 03:45	Received: 10/06/16 10:15	Matrix: Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS, DW		Analytical Method: EPA 200.8						
Lead	1.6	ug/L	0.10	1		10/10/16 14:24	7439-92-1	
Sample: ROOM 140/CS		Lab ID: 10365060066	Collected: 09/27/16 03:46	Received: 10/06/16 10:15	Matrix: Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS, DW		Analytical Method: EPA 200.8						
Lead	0.84	ug/L	0.10	1		10/10/16 14:28	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

ANALYTICAL RESULTS

Project: Han CSD - Fairly Elementary

Pace Project No.: 10365060

Sample: ROOM 140/BS		Lab ID: 10365060067	Collected: 09/27/16 03:46	Received: 10/06/16 10:15	Matrix: Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS, DW		Analytical Method: EPA 200.8						
Lead	0.70	ug/L	0.10	1		10/10/16 14:30	7439-92-1	
Sample: ROOM 139/CS		Lab ID: 10365060068	Collected: 09/27/16 03:47	Received: 10/06/16 10:15	Matrix: Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS, DW		Analytical Method: EPA 200.8						
Lead	1.1	ug/L	0.10	1		10/10/16 14:31	7439-92-1	
Sample: ROOM 139/BS		Lab ID: 10365060069	Collected: 09/27/16 03:47	Received: 10/06/16 10:15	Matrix: Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS, DW		Analytical Method: EPA 200.8						
Lead	1.0	ug/L	0.10	1		10/10/16 14:33	7439-92-1	
Sample: ROOM 141/CS		Lab ID: 10365060070	Collected: 09/27/16 03:50	Received: 10/06/16 10:15	Matrix: Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS, DW		Analytical Method: EPA 200.8						
Lead	0.91	ug/L	0.10	1		10/10/16 14:34	7439-92-1	
Sample: ROOM 141/BS		Lab ID: 10365060071	Collected: 09/27/16 03:50	Received: 10/06/16 10:15	Matrix: Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS, DW		Analytical Method: EPA 200.8						
Lead	0.79	ug/L	0.10	1		10/10/16 14:36	7439-92-1	
Sample: ROOM 201/CS		Lab ID: 10365060072	Collected: 09/27/16 03:55	Received: 10/06/16 10:15	Matrix: Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS, DW		Analytical Method: EPA 200.8						
Lead	0.76	ug/L	0.10	1		10/10/16 14:39	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

ANALYTICAL RESULTS

Project: Han CSD - Fairly Elementary

Pace Project No.: 10365060

Sample: ROOM 202/CS		Lab ID: 10365060073	Collected: 09/27/16 03:55	Received: 10/06/16 10:15	Matrix: Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS, DW		Analytical Method: EPA 200.8						
Lead	1.4	ug/L	0.10	1		10/10/16 14:57	7439-92-1	
Sample: ROOM 203/CS		Lab ID: 10365060074	Collected: 09/27/16 03:56	Received: 10/06/16 10:15	Matrix: Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS, DW		Analytical Method: EPA 200.8						
Lead	1.0	ug/L	0.10	1		10/10/16 14:58	7439-92-1	
Sample: ROOM 204/CS		Lab ID: 10365060075	Collected: 09/27/16 03:56	Received: 10/06/16 10:15	Matrix: Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS, DW		Analytical Method: EPA 200.8						
Lead	0.88	ug/L	0.10	1		10/10/16 15:06	7439-92-1	
Sample: ROOM 205/CS		Lab ID: 10365060076	Collected: 09/27/16 03:57	Received: 10/06/16 10:15	Matrix: Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS, DW		Analytical Method: EPA 200.8						
Lead	1.1	ug/L	0.10	1		10/10/16 15:07	7439-92-1	
Sample: 2NDFLR/BOYS ROOM/SINK1		Lab ID: 10365060077	Collected: 09/27/16 03:59	Received: 10/06/16 10:15	Matrix: Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS, DW		Analytical Method: EPA 200.8						
Lead	0.34	ug/L	0.10	1		10/10/16 15:09	7439-92-1	
Sample: 2NDFLR/BOYS ROOM/SINK2		Lab ID: 10365060078	Collected: 09/27/16 04:00	Received: 10/06/16 10:15	Matrix: Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS, DW		Analytical Method: EPA 200.8						
Lead	0.32	ug/L	0.10	1		10/10/16 15:10	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

ANALYTICAL RESULTS

Project: Han CSD - Fairly Elementary

Pace Project No.: 10365060

Sample: 2NDFLR/BOYS ROOM/SINK3		Lab ID: 10365060079	Collected: 09/27/16 04:01	Received: 10/06/16 10:15	Matrix: Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS, DW		Analytical Method: EPA 200.8						
Lead	0.57	ug/L	0.10	1		10/10/16 15:12	7439-92-1	
Sample: 2NDFLR/BOYS ROOM/SINK4		Lab ID: 10365060080	Collected: 09/27/16 04:02	Received: 10/06/16 10:15	Matrix: Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS, DW		Analytical Method: EPA 200.8						
Lead	0.53	ug/L	0.10	1		10/10/16 15:13	7439-92-1	
Sample: 2NDFLR/BOYS ROOM/SINK5		Lab ID: 10365060081	Collected: 09/27/16 04:02	Received: 10/06/16 10:15	Matrix: Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS, DW		Analytical Method: EPA 200.8						
Lead	0.39	ug/L	0.10	1		10/12/16 20:21	7439-92-1	
Sample: 2NDFLR/BOYS ROOM/SINK6		Lab ID: 10365060082	Collected: 09/27/16 04:03	Received: 10/06/16 10:15	Matrix: Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS, DW		Analytical Method: EPA 200.8						
Lead	0.56	ug/L	0.10	1		10/12/16 16:57	7439-92-1	
Sample: 2NDFLR/FACULTY BATH/LEFT SINK		Lab ID: 10365060083	Collected: 09/27/16 04:06	Received: 10/06/16 10:15	Matrix: Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS, DW		Analytical Method: EPA 200.8						
Lead	1.1	ug/L	0.10	1		10/12/16 17:07	7439-92-1	
Sample: 2NDFLR/FACULTY BATH/RIGHT SINK		Lab ID: 10365060084	Collected: 09/27/16 04:06	Received: 10/06/16 10:15	Matrix: Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS, DW		Analytical Method: EPA 200.8						
Lead	1.4	ug/L	0.10	1		10/12/16 17:09	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

ANALYTICAL RESULTS

Project: Han CSD - Fairly Elementary

Pace Project No.: 10365060

Sample: ROOM 206/CS		Lab ID: 10365060085	Collected: 09/27/16 04:08	Received: 10/06/16 10:15	Matrix: Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS, DW		Analytical Method: EPA 200.8						
Lead	0.61	ug/L	0.10	1		10/12/16 17:10	7439-92-1	
Sample: ROOM 207/CS		Lab ID: 10365060086	Collected: 09/27/16 04:09	Received: 10/06/16 10:15	Matrix: Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS, DW		Analytical Method: EPA 200.8						
Lead	1.1	ug/L	0.10	1		10/11/16 18:35	7439-92-1	
Sample: 2ND FLR/WOMENS FACULTY/LEFT		Lab ID: 10365060087	Collected: 09/27/16 04:10	Received: 10/06/16 10:15	Matrix: Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS, DW		Analytical Method: EPA 200.8						
Lead	0.64	ug/L	0.10	1		10/11/16 18:37	7439-92-1	
Sample: 2ND FLR/WOMENS FACULTY/RIGHT		Lab ID: 10365060088	Collected: 09/27/16 04:11	Received: 10/06/16 10:15	Matrix: Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS, DW		Analytical Method: EPA 200.8						
Lead	0.76	ug/L	0.10	1		10/11/16 18:38	7439-92-1	
Sample: 2ND FLR/GIRLS ROOM/SINK 1		Lab ID: 10365060089	Collected: 09/27/16 04:12	Received: 10/06/16 10:15	Matrix: Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS, DW		Analytical Method: EPA 200.8						
Lead	0.37	ug/L	0.10	1		10/11/16 18:39	7439-92-1	
Sample: 2ND FLR/GIRLS ROOM/SINK 2		Lab ID: 10365060090	Collected: 09/27/16 04:13	Received: 10/06/16 10:15	Matrix: Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS, DW		Analytical Method: EPA 200.8						
Lead	0.43	ug/L	0.10	1		10/11/16 18:41	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

ANALYTICAL RESULTS

Project: Han CSD - Fairly Elementary

Pace Project No.: 10365060

Sample:	Lab ID:	Collected:	Received:	Matrix:				
2ND FLR/GIRLS ROOM/SINK 3	10365060091	09/27/16 04:14	10/06/16 10:15	Water				
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual

200.8 MET ICPMS, DW	Analytical Method: EPA 200.8							
Lead	0.90	ug/L	0.10	1		10/11/16 18:42	7439-92-1	

Sample:	Lab ID:	Collected:	Received:	Matrix:				
2ND FLR/GIRLS ROOM/SINK 4	10365060092	09/27/16 04:15	10/06/16 10:15	Water				
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual

200.8 MET ICPMS, DW	Analytical Method: EPA 200.8							
Lead	0.75	ug/L	0.10	1		10/11/16 18:45	7439-92-1	

Sample:	Lab ID:	Collected:	Received:	Matrix:				
2ND FLR/GIRLS ROOM/SINK 5	10365060093	09/27/16 04:16	10/06/16 10:15	Water				
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual

200.8 MET ICPMS	Analytical Method: EPA 200.8 Preparation Method: EPA 200.8							
Lead	0.60	ug/L	0.10	1	10/12/16 09:30	10/12/16 17:14	7439-92-1	

Sample:	Lab ID:	Collected:	Received:	Matrix:				
2ND FLR/GIRLS ROOM/SINK 6	10365060094	09/27/16 04:17	10/06/16 10:15	Water				
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual

200.8 MET ICPMS, DW	Analytical Method: EPA 200.8							
Lead	0.45	ug/L	0.10	1		10/11/16 18:46	7439-92-1	

Sample:	Lab ID:	Collected:	Received:	Matrix:				
ROOM 209/CS	10365060095	09/27/16 04:17	10/06/16 10:15	Water				
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual

200.8 MET ICPMS, DW	Analytical Method: EPA 200.8							
Lead	1.4	ug/L	0.10	1		10/11/16 18:47	7439-92-1	

Sample:	Lab ID:	Collected:	Received:	Matrix:				
ROOM 208/CS	10365060096	09/27/16 04:18	10/06/16 10:15	Water				
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual

200.8 MET ICPMS, DW	Analytical Method: EPA 200.8							
Lead	1.3	ug/L	0.10	1		10/11/16 18:52	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full, without the written consent of Pace Analytical Services, LLC.

ANALYTICAL RESULTS

Project: Han CSD - Fairly Elementary

Pace Project No.: 10365060

Sample: ROOM 210/CS		Lab ID: 10365060097	Collected: 09/27/16 04:20	Received: 10/06/16 10:15	Matrix: Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS, DW		Analytical Method: EPA 200.8						
Lead	1.3	ug/L	0.10	1		10/11/16 18:53	7439-92-1	
Sample: ROOM 212/CS		Lab ID: 10365060098	Collected: 09/27/16 04:21	Received: 10/06/16 10:15	Matrix: Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS, DW		Analytical Method: EPA 200.8						
Lead	1.0	ug/L	0.10	1		10/11/16 18:54	7439-92-1	
Sample: ROOM 211/CS		Lab ID: 10365060099	Collected: 09/27/16 04:21	Received: 10/06/16 10:15	Matrix: Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS, DW		Analytical Method: EPA 200.8						
Lead	2.2	ug/L	0.10	1		10/11/16 18:56	7439-92-1	
Sample: ROOM 213/CS		Lab ID: 10365060100	Collected: 09/27/16 04:22	Received: 10/06/16 10:15	Matrix: Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS, DW		Analytical Method: EPA 200.8						
Lead	2.4	ug/L	0.10	1		10/11/16 18:57	7439-92-1	
Sample: ROOM 215/CS		Lab ID: 10365060101	Collected: 09/27/16 04:24	Received: 10/06/16 10:15	Matrix: Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS, DW		Analytical Method: EPA 200.8						
Lead	1.4	ug/L	0.10	1		10/11/16 17:30	7439-92-1	
Sample: ROOM 214/CS		Lab ID: 10365060102	Collected: 09/27/16 04:25	Received: 10/06/16 10:15	Matrix: Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS, DW		Analytical Method: EPA 200.8						
Lead	0.70	ug/L	0.10	1		10/11/16 17:35	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

ANALYTICAL RESULTS

Project: Han CSD - Fairly Elementary

Pace Project No.: 10365060

Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
Sample: ROOM 216/CS		Lab ID: 10365060103		Collected: 09/27/16 04:25	Received: 10/06/16 10:15	Matrix: Water		
200.8 MET ICPMS, DW	Analytical Method: EPA 200.8							
Lead	0.63	ug/L	0.10	1		10/11/16 17:37	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

QUALITY CONTROL DATA

Project: Han CSD - Fairly Elementary
Pace Project No.: 10365060

QC Batch: 439834 Analysis Method: EPA 200.8
QC Batch Method: EPA 200.8 Analysis Description: ICPMS Metals, Drinking Water
Associated Lab Samples: 10365060001, 10365060002, 10365060003, 10365060004, 10365060005, 10365060006, 10365060007, 10365060008, 10365060009, 10365060010, 10365060011, 10365060012, 10365060013, 10365060014, 10365060015, 10365060016, 10365060017, 10365060018, 10365060019, 10365060020

METHOD BLANK: 2391326 Matrix: Water
Associated Lab Samples: 10365060001, 10365060002, 10365060003, 10365060004, 10365060005, 10365060006, 10365060007, 10365060008, 10365060009, 10365060010, 10365060011, 10365060012, 10365060013, 10365060014, 10365060015, 10365060016, 10365060017, 10365060018, 10365060019, 10365060020

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Lead	ug/L	ND	0.10	10/11/16 17:16	

LABORATORY CONTROL SAMPLE: 2391327

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	100	102	102	85-115	

MATRIX SPIKE & MATRIX SPIKE DUPLICATE: 2395632 2395633

Parameter	Units	10365060001 Result	MS Spike Conc.	MSD Spike Conc.	MS Result	MSD Result	MS % Rec	MSD % Rec	% Rec Limits	RPD	Max RPD	Qual
Lead	ug/L	10.9	100	100	111	111	100	100	70-130	0	20	

MATRIX SPIKE SAMPLE: 2395634

Parameter	Units	10365060011 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	0.37	100	100	100	70-130	

Results presented on this page are in the units indicated by the "Units" column except where an alternate unit is presented to the right of the result.

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

QUALITY CONTROL DATA

Project: Han CSD - Fairly Elementary

Pace Project No.: 10365060

QC Batch:	439835	Analysis Method:	EPA 200.8
QC Batch Method:	EPA 200.8	Analysis Description:	ICPMS Metals, Drinking Water
Associated Lab Samples:	10365060021, 10365060022, 10365060023, 10365060024, 10365060025, 10365060026, 10365060027, 10365060029, 10365060030, 10365060031, 10365060032, 10365060033, 10365060035, 10365060036, 10365060037, 10365060038, 10365060039, 10365060040		

METHOD BLANK:	2391329	Matrix:	Water
Associated Lab Samples:	10365060021, 10365060022, 10365060023, 10365060024, 10365060025, 10365060026, 10365060027, 10365060029, 10365060030, 10365060031, 10365060032, 10365060033, 10365060035, 10365060036, 10365060037, 10365060038, 10365060039, 10365060040		

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Lead	ug/L	ND	0.10	10/11/16 19:42	

LABORATORY CONTROL SAMPLE: 2391330

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	100	102	102	85-115	

MATRIX SPIKE & MATRIX SPIKE DUPLICATE: 2395596 2395597

Parameter	Units	10365060021 Result	MS Spike Conc.	MSD Spike Conc.	MS Result	MSD Result	MS % Rec	MSD % Rec	% Rec Limits	RPD	Max RPD	Qual
Lead	ug/L	0.39	100	100	93.9	94.0	94	94	70-130	0	20	

MATRIX SPIKE SAMPLE: 2395598

Parameter	Units	10365060032 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	9.1	100	104	95	70-130	

Results presented on this page are in the units indicated by the "Units" column except where an alternate unit is presented to the right of the result.

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full, without the written consent of Pace Analytical Services, LLC.

QUALITY CONTROL DATA

Project: Han CSD - Fairly Elementary

Pace Project No.: 10365060

QC Batch: 439836 Analysis Method: EPA 200.8
 QC Batch Method: EPA 200.8 Analysis Description: ICPMS Metals, Drinking Water
 Associated Lab Samples: 10365060041, 10365060042, 10365060043, 10365060044, 10365060045, 10365060046, 10365060047, 10365060048, 10365060049, 10365060050, 10365060051, 10365060052, 10365060053, 10365060054, 10365060055, 10365060056, 10365060057, 10365060058, 10365060059, 10365060060

METHOD BLANK: 2391335 Matrix: Water
 Associated Lab Samples: 10365060041, 10365060042, 10365060043, 10365060044, 10365060045, 10365060046, 10365060047, 10365060048, 10365060049, 10365060050, 10365060051, 10365060052, 10365060053, 10365060054, 10365060055, 10365060056, 10365060057, 10365060058, 10365060059, 10365060060

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Lead	ug/L	ND	0.10	10/11/16 18:58	

LABORATORY CONTROL SAMPLE: 2391336

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	100	101	101	85-115	

MATRIX SPIKE & MATRIX SPIKE DUPLICATE: 2395442 2395443

Parameter	Units	10365060041 Result	MS Spike Conc.	MSD Spike Conc.	MS Result	MSD Result	MS % Rec	MSD % Rec	% Rec Limits	RPD	Max RPD	Qual
Lead	ug/L	0.87	100	100	98.0	98.5	97	98	70-130	1	20	

MATRIX SPIKE SAMPLE: 2395444

Parameter	Units	10365060051 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	1.4	100	99.2	98	70-130	

Results presented on this page are in the units indicated by the "Units" column except where an alternate unit is presented to the right of the result.

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full, without the written consent of Pace Analytical Services, LLC.

QUALITY CONTROL DATA

Project: Han CSD - Fairly Elementary
Pace Project No.: 10365060

QC Batch: 439837 Analysis Method: EPA 200.8
QC Batch Method: EPA 200.8 Analysis Description: ICPMS Metals, Drinking Water
Associated Lab Samples: 10365060061, 10365060062, 10365060063, 10365060064, 10365060065, 10365060066, 10365060067, 10365060068, 10365060069, 10365060070, 10365060071, 10365060072, 10365060073, 10365060074, 10365060075, 10365060076, 10365060077, 10365060078, 10365060079, 10365060080

METHOD BLANK: 2391338 Matrix: Water
Associated Lab Samples: 10365060061, 10365060062, 10365060063, 10365060064, 10365060065, 10365060066, 10365060067, 10365060068, 10365060069, 10365060070, 10365060071, 10365060072, 10365060073, 10365060074, 10365060075, 10365060076, 10365060077, 10365060078, 10365060079, 10365060080

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Lead	ug/L	ND	0.10	10/10/16 14:10	

LABORATORY CONTROL SAMPLE: 2391339

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	100	93.9	94	85-115	

MATRIX SPIKE & MATRIX SPIKE DUPLICATE: 2393626 2393627

Parameter	Units	10365060061 Result	MS Spike Conc.	MSD Spike Conc.	MS Result	MSD Result	MS % Rec	MSD % Rec	% Rec Limits	RPD	Max RPD	Qual
Lead	ug/L	0.64	100	100	98.3	97.0	98	96	70-130	1	20	

MATRIX SPIKE SAMPLE: 2393628

Parameter	Units	10365060071 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	0.79	100	102	101	70-130	

Results presented on this page are in the units indicated by the "Units" column except where an alternate unit is presented to the right of the result.

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full, without the written consent of Pace Analytical Services, LLC.

QUALITY CONTROL DATA

Project: Han CSD - Fairly Elementary

Pace Project No.: 10365060

QC Batch:	439838	Analysis Method:	EPA 200.8
QC Batch Method:	EPA 200.8	Analysis Description:	ICPMS Metals, Drinking Water
Associated Lab Samples:	10365060081, 10365060082, 10365060083, 10365060084, 10365060085, 10365060086, 10365060087, 10365060088, 10365060089, 10365060090, 10365060091, 10365060092, 10365060094, 10365060095, 10365060096, 10365060097, 10365060098, 10365060099, 10365060100		

METHOD BLANK:	2391341	Matrix:	Water
Associated Lab Samples:	10365060081, 10365060082, 10365060083, 10365060084, 10365060085, 10365060086, 10365060087, 10365060088, 10365060089, 10365060090, 10365060091, 10365060092, 10365060094, 10365060095, 10365060096, 10365060097, 10365060098, 10365060099, 10365060100		

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Lead	ug/L	ND	0.10	10/12/16 20:19	

LABORATORY CONTROL SAMPLE: 2391342

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	100	101	101	85-115	

MATRIX SPIKE & MATRIX SPIKE DUPLICATE: 2395415 2395416

Parameter	Units	10365060081 Result	MS Spike Conc.	MSD Spike Conc.	MS Result	MSD Result	MS % Rec	MSD % Rec	% Rec Limits	RPD	Max RPD	Qual
Lead	ug/L	0.39	100	100	94.3	96.9	94	96	70-130	3	20	

MATRIX SPIKE SAMPLE: 2395417

Parameter	Units	10365060091 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	0.90	100	98.5	98	70-130	

Results presented on this page are in the units indicated by the "Units" column except where an alternate unit is presented to the right of the result.

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full, without the written consent of Pace Analytical Services, LLC.

QUALITY CONTROL DATA

Project: Han CSD - Fairly Elementary

Pace Project No.: 10365060

QC Batch: 439839

Analysis Method: EPA 200.8

QC Batch Method: EPA 200.8

Analysis Description: ICPMS Metals, Drinking Water

Associated Lab Samples: 10365060101, 10365060102, 10365060103

METHOD BLANK: 2391349

Matrix: Water

Associated Lab Samples: 10365060101, 10365060102, 10365060103

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Lead	ug/L	ND	0.10	10/11/16 17:27	

LABORATORY CONTROL SAMPLE: 2391350

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	100	100	100	85-115	

MATRIX SPIKE & MATRIX SPIKE DUPLICATE: 2395383 2395384

Parameter	Units	10365060101		2395383		2395384		% Rec Limits	RPD	Max RPD	Qual	
		MS Result	MSD Result	MS Spike Conc.	MSD Spike Conc.	MS Result	MSD Result					
Lead	ug/L	1.4	1.4	100	100	100	101	99	100	70-130	1	20

MATRIX SPIKE SAMPLE: 2395385

Parameter	Units	10365068008 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	15.1	100	113	98	70-130	

Results presented on this page are in the units indicated by the "Units" column except where an alternate unit is presented to the right of the result.

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

QUALITY CONTROL DATA

Project: Han CSD - Fairly Elementary

Pace Project No.: 10365060

QC Batch: 440200 Analysis Method: EPA 200.8

QC Batch Method: EPA 200.8 Analysis Description: 200.8 MET

Associated Lab Samples: 10365060028, 10365060034, 10365060093

METHOD BLANK: 2393729 Matrix: Water

Associated Lab Samples: 10365060028, 10365060034, 10365060093

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Lead	ug/L	ND	0.10	10/12/16 16:47	

LABORATORY CONTROL SAMPLE: 2393730

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	100	101	101	85-115	

MATRIX SPIKE & MATRIX SPIKE DUPLICATE: 2393731 2393732

Parameter	Units	10364697007		2393731		2393732		% Rec Limits	RPD	Max RPD	Qual	
		MS Result	MSD Result	MS Spike Conc.	MSD Spike Conc.	MS Result	MSD Result					MS % Rec
Lead	ug/L	4.2	4.2	100	100	97.0	102	93	98	70-130	5	20

MATRIX SPIKE SAMPLE: 2393733

Parameter	Units	10365090028 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	30.1	100	129	99	70-130	

Results presented on this page are in the units indicated by the "Units" column except where an alternate unit is presented to the right of the result.

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

QUALIFIERS

Project: Han CSD - Fairly Elementary

Pace Project No.: 10365060

DEFINITIONS

DF - Dilution Factor, if reported, represents the factor applied to the reported data due to dilution of the sample aliquot.

ND - Not Detected at or above adjusted reporting limit.

J - Estimated concentration above the adjusted method detection limit and below the adjusted reporting limit.

MDL - Adjusted Method Detection Limit.

PQL - Practical Quantitation Limit.

RL - Reporting Limit.

S - Surrogate

1,2-Diphenylhydrazine decomposes to and cannot be separated from Azobenzene using Method 8270. The result for each analyte is a combined concentration.

Consistent with EPA guidelines, unrounded data are displayed and have been used to calculate % recovery and RPD values.

LCS(D) - Laboratory Control Sample (Duplicate)

MS(D) - Matrix Spike (Duplicate)

DUP - Sample Duplicate

RPD - Relative Percent Difference

NC - Not Calculable.

SG - Silica Gel - Clean-Up

U - Indicates the compound was analyzed for, but not detected.

N-Nitrosodiphenylamine decomposes and cannot be separated from Diphenylamine using Method 8270. The result reported for each analyte is a combined concentration.

Pace Analytical is TNI accredited. Contact your Pace PM for the current list of accredited analytes.

TNI - The NELAC Institute.

LABORATORIES

PASI-M Pace Analytical Services - Minneapolis

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

QUALITY CONTROL DATA CROSS REFERENCE TABLE

Project: Han CSD - Fairly Elementary

Pace Project No.: 10365060

Lab ID	Sample ID	QC Batch Method	QC Batch	Analytical Method	Analytical Batch
10365060001	Kitchen/Disposal Sprayer	EPA 200.8	439834		
10365060002	Cafeteria/Girls Room	EPA 200.8	439834		
10365060003	Cafeteria/Boys Room	EPA 200.8	439834		
10365060004	Room 040/ CS	EPA 200.8	439834		
10365060005	Room 030/ CS	EPA 200.8	439834		
10365060006	Room 010/ CS	EPA 200.8	439834		
10365060007	Room 010/ BS	EPA 200.8	439834		
10365060008	Room 020/ CS	EPA 200.8	439834		
10365060009	Room 020/ BS	EPA 200.8	439834		
10365060010	MAIN ENT./LADIES RM/LEFT SINK	EPA 200.8	439834		
10365060011	MAIN ENT./LADIES RM/RIGHT SINK	EPA 200.8	439834		
10365060012	MAIN ENT./MENS RM/LEFT SINK	EPA 200.8	439834		
10365060013	MAIN ENT./MENS RM/RIGHT SINK	EPA 200.8	439834		
10365060014	ROOM 107/CS	EPA 200.8	439834		
10365060015	ROOM 105/CS	EPA 200.8	439834		
10365060016	ROOM 106/CS	EPA 200.8	439834		
10365060017	ROOM 103/CS	EPA 200.8	439834		
10365060018	ROOM 104/CS	EPA 200.8	439834		
10365060019	ROOM 102/CS	EPA 200.8	439834		
10365060020	ROOM 102/BS	EPA 200.8	439834		
10365060021	ROOM 101/CS	EPA 200.8	439835		
10365060022	ROOM 101/BS	EPA 200.8	439835		
10365060023	TEACHERS LOUNGE	EPA 200.8	439835		
10365060024	RESPECT RD/MENS RM SINK	EPA 200.8	439835		
10365060025	RESPECT RD/LADIES RM SINK	EPA 200.8	439835		
10365060026	ROOM 126/CS	EPA 200.8	439835		
10365060027	ROOM 121/CS	EPA 200.8	439835		
10365060029	ROOM 122/CS	EPA 200.8	439835		
10365060030	ROOM 125/CS	EPA 200.8	439835		
10365060031	ROOM 124/CS	EPA 200.8	439835		
10365060032	ROOM 126/CS	EPA 200.8	439835		
10365060033	ROOM 127/CS	EPA 200.8	439835		
10365060035	ROOM 130/BS	EPA 200.8	439835		
10365060036	ROOM 131/BS-LEFT	EPA 200.8	439835		
10365060037	ROOM 131/BS-RIGHT	EPA 200.8	439835		
10365060038	BOYS LOCKER ROOM/LEFT SINK	EPA 200.8	439835		
10365060039	BOYS LOCKER ROOM/RIGHT SINK	EPA 200.8	439835		
10365060040	BOYS GYM/OFFICE BATH	EPA 200.8	439835		
10365060041	ROOM 132/BS	EPA 200.8	439836		
10365060042	MENS RM/ACROSS132/SINK1	EPA 200.8	439836		
10365060043	MENS RM/ACROSS132/SINK2	EPA 200.8	439836		
10365060044	MENS RM/ACROSS132/SINK3	EPA 200.8	439836		
10365060045	MENS RM/ACROSS132/SINK4	EPA 200.8	439836		
10365060046	LadiesRM/ACROSS132/SINK1	EPA 200.8	439836		
10365060047	LadiesRM/ACROSS132/SINK2	EPA 200.8	439836		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

QUALITY CONTROL DATA CROSS REFERENCE TABLE

Project: Han CSD - Fairly Elementary
Pace Project No.: 10365060

Lab ID	Sample ID	QC Batch Method	QC Batch	Analytical Method	Analytical Batch
10365060048	LadiesRM/ACROSS132/SINK3	EPA 200.8	439836		
10365060049	LadiesRM/ACROSS132/SINK4	EPA 200.8	439836		
10365060050	ROOM 133/LEFT SINK	EPA 200.8	439836		
10365060051	ROOM 133/RIGHT SINK	EPA 200.8	439836		
10365060052	ROOM 133/BS	EPA 200.8	439836		
10365060053	GIRLS LOCKER ROOM/LEFT SINK	EPA 200.8	439836		
10365060054	GIRLS LOCKER ROOM/RIGHT SINK	EPA 200.8	439836		
10365060055	GIRLS COACHES BATH	EPA 200.8	439836		
10365060056	ROOM 134/CS	EPA 200.8	439836		
10365060057	ROOM 134/BS	EPA 200.8	439836		
10365060058	ROOM 136/CS	EPA 200.8	439836		
10365060059	ROOM 136/BS	EPA 200.8	439836		
10365060060	ROOM 135/BS	EPA 200.8	439836		
10365060061	ROOM 135/CS	EPA 200.8	439837		
10365060062	ROOM 137/CS	EPA 200.8	439837		
10365060063	ROOM 137/BS	EPA 200.8	439837		
10365060064	ROOM 138/CS	EPA 200.8	439837		
10365060065	ROOM 138/BS	EPA 200.8	439837		
10365060066	ROOM 140/CS	EPA 200.8	439837		
10365060067	ROOM 140/BS	EPA 200.8	439837		
10365060068	ROOM 139/CS	EPA 200.8	439837		
10365060069	ROOM 139/BS	EPA 200.8	439837		
10365060070	ROOM 141/CS	EPA 200.8	439837		
10365060071	ROOM 141/BS	EPA 200.8	439837		
10365060072	ROOM 201/CS	EPA 200.8	439837		
10365060073	ROOM 202/CS	EPA 200.8	439837		
10365060074	ROOM 203/CS	EPA 200.8	439837		
10365060075	ROOM 204/CS	EPA 200.8	439837		
10365060076	ROOM 205/CS	EPA 200.8	439837		
10365060077	2NDFLR/BOYS ROOM/SINK1	EPA 200.8	439837		
10365060078	2NDFLR/BOYS ROOM/SINK2	EPA 200.8	439837		
10365060079	2NDFLR/BOYS ROOM/SINK3	EPA 200.8	439837		
10365060080	2NDFLR/BOYS ROOM/SINK4	EPA 200.8	439837		
10365060081	2NDFLR/BOYS ROOM/SINK5	EPA 200.8	439838		
10365060082	2NDFLR/BOYS ROOM/SINK6	EPA 200.8	439838		
10365060083	2NDFLR/FACULTY BATH/LEFT SINK	EPA 200.8	439838		
10365060084	2NDFLR/FACULTY BATH/RIGHT SINK	EPA 200.8	439838		
10365060085	ROOM 206/CS	EPA 200.8	439838		
10365060086	ROOM 207/CS	EPA 200.8	439838		
10365060087	2ND FLR/WOMENS FACULTY/LEFT	EPA 200.8	439838		
10365060088	2ND FLR/WOMENS FACULTY/RIGHT	EPA 200.8	439838		
10365060089	2ND FLR/GIRLS ROOM/SINK 1	EPA 200.8	439838		
10365060090	2ND FLR/GIRLS ROOM/SINK 2	EPA 200.8	439838		
10365060091	2ND FLR/GIRLS ROOM/SINK 3	EPA 200.8	439838		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

QUALITY CONTROL DATA CROSS REFERENCE TABLE

Project: Han CSD - Fairly Elementary

Pace Project No.: 10365060

Lab ID	Sample ID	QC Batch Method	QC Batch	Analytical Method	Analytical Batch
10365060092	2ND FLR/GIRLS ROOM/SINK 4	EPA 200.8	439838		
10365060094	2ND FLR/GIRLS ROOM/SINK 6	EPA 200.8	439838		
10365060095	ROOM 209/CS	EPA 200.8	439838		
10365060096	ROOM 208/CS	EPA 200.8	439838		
10365060097	ROOM 210/CS	EPA 200.8	439838		
10365060098	ROOM 212/CS	EPA 200.8	439838		
10365060099	ROOM 211/CS	EPA 200.8	439838		
10365060100	ROOM 213/CS	EPA 200.8	439838		
10365060101	ROOM 215/CS	EPA 200.8	439839		
10365060102	ROOM 214/CS	EPA 200.8	439839		
10365060103	ROOM 216/CS	EPA 200.8	439839		
10365060028	ROOM 123/CS	EPA 200.8	440200	EPA 200.8	440647
10365060034	ROOM 130/CS	EPA 200.8	440200	EPA 200.8	440647
10365060093	2ND FLR/GIRLS ROOM/SINK 5	EPA 200.8	440200	EPA 200.8	440647

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

CHAIN-OF-CUSTODY / Analytical Request Document

The Chain-of-Custody is a LEGAL DOCUMENT. All relevant fields must be completed accurately.



10365660
Page: 1 of 9

Section A Required Client Information:	Section B Required Project Information:	Section C Invoice Information:
Company: Hannibal CSD	Report To:	Attention:
Address: 928 Cayuga St	Copy To:	Company Name:
Hannibal, NY 13074	Purchase Order No.:	Address:
Email To: nhennner@hannibalcsd.org	Project Name: Fairly Elementary	Pace Quote Reference:
Phone:	Project Number:	Pace Project Manager:
Requested Due Date/TAT:		Pace Profile #:

REGULATORY AGENCY

NPDES GROUND WATER DRINKING WATER

UST RCRA OTHER

SITE LOCATION

GA IL IN MI NC
 OH SC WI OTHER_NY

# ITEM	Section D Required Client Information	Valid Matrix Codes	MATRIX CODE	COLLECTED		SAMPLE TYPE	G-GRAB C-COMP	# OF CONTAINERS	PRESERVATIVES	Requester Name (Y/N)	Pace Project No. Lab I.D.
				COMPOSITE START	COMPOSITE END/GRAB						
1	KITCHEN / DISPOSAL SINK	DW	DW	9/27	2:31	DW	G	1	Unpreserved	X	01
2	CAFETERIA / GIRLS ROOM	DW	DW		2:34	DW	G	1	H ₂ SO ₄		02
3	CAFETERIA / BOYS ROOM	DW	DW		2:34	DW	G	1	HNO ₃		03
4	ROOM 040 / BS	DW	DW		2:36	DW	G	1	HNO ₃		04
5	ROOM 030 / CS	DW	DW		2:37	DW	G	1	HNO ₃		05
6	ROOM 010 / CS	DW	DW		2:37	DW	G	1	HNO ₃		06
7	ROOM 010 / BS	DW	DW		2:38	DW	G	1	HNO ₃		07
8	ROOM 020 / CS	DW	DW		2:38	DW	G	1	HNO ₃		08
9	ROOM 020 / BS	DW	DW		2:40	DW	G	1	HNO ₃		09
10	MAIN ENT. / LADIES RM / LEFT SINK	DW	DW		2:43	DW	G	1	HNO ₃		010
11	MAIN ENT. / LADIES RM / RIGHT SINK	DW	DW		2:44	DW	G	1	HNO ₃		011
12	MAIN ENT. / MENS RM / LEFT SINK	DW	DW		2:45	DW	G	1	HNO ₃		012

ADDITIONAL COMMENTS	RELINQUISHED BY / AFFILIATION	DATE	TIME	ACCEPTED BY / AFFILIATION	DATE	TIME	SAMPLE CONDITIONS
CS = CLASSROOM SINK BS = BATHROOM SINK	James PACE	10/3/16	17:00	James PACE	10/6/16	10:15	Y/N Y/N Y/N Y/N Y/N Y/N Y/N Y/N Y/N Y/N Y/N Y/N

SAMPLER NAME AND SIGNATURE

PRINT Name of SAMPLER: James Murphy - PACE Analytical

SIGNATURE of SAMPLER: *James Murphy*

DATE Signed (MM/DD/YY): 9/27/16

CHAIN-OF-CUSTODY / Analytical Request Document

The Chain-of-Custody is a LEGAL DOCUMENT. All relevant fields must be completed accurately.



Section A

Required Client Information:

Company: **Hannibal CSD**
 Address: **928 Cayuga St**
Hannibal, NY 13074
 Email To: **nhenner@hannibalcso.org**
 Phone: _____
 Requested Due Date/TAT: _____

Section B

Report To:

Copy To: _____
 Address: _____
 Pace Quote Reference: _____
 Pace Project Manager: _____
 Project Name: **Fairly Elementary**
 Project Number: _____

Section C

Invoice Information:

Attention: _____
 Company Name: _____
 Address: _____
 Pace Quote Reference: _____
 Pace Project Manager: _____
 Pace Profile #: _____

REGULATORY AGENCY

NPDES GROUND WATER DRINKING WATER
 UST RCRA OTHER _____

SITE GA IL IN MI NC
LOCATION OH SC WI OTHER NY

#	ITEM	Section D Required Client Information	Valid Matrix Codes	MATRIX	CODE	COLLECTED		# OF CONTAINERS	PRESERVATIVES	Requester Ant	Filtered (Y/N)	Pace Project No. Lab I.D.			
						COMPOSITE START	COMPOSITE END/GRAB								
		SAMPLE ID (A-Z, 0-9 / -)	DW WW B OL WP OT TS	DRINKING WATER WASTE WATER PRODUCT OIL WIRE OTHER TISSUE	G-GRAB C-COMP	DATE	TIME		H ₂ SO ₄	HNO ₃	HCl	NaOH	Na ₂ SO ₄	Methanol	Na ₂ SO ₄
1		MAIN ENT./MENS RM /RIGHT SINK	DW		G	9/27	2:45	1		X		013			
2		Room 107/CS	DW		G		2:46	1				014			
3		Room 105/CS	DW		G		2:46	1				015			
4		Room 106/CS	DW		G		2:47	1				016			
5		Room 103/CS	DW		G		2:47	1				017			
6		Room 104/CS	DW		G		2:48	1				018			
7		Room 102/CS	DW		G		2:50	1				019			
8		Room 102/BS	DW		G		2:50	1				020			
9		Room 101/CS	DW		G		2:51	1				021			
10		Room 101/BS	DW		G		2:52	1				022			
11		TEACHERS LOUNGE	DW		G		2:52	1				023			
12		RESTFRD/MENS RM SINK	DW		G		2:52	1				024			

ADDITIONAL COMMENTS	RELINQUISHED BY / AFFILIATION	DATE	TIME	ACCEPTED BY / AFFILIATION	DATE	TIME	SAMPLE CONDITIONS
	<i>James Murphy</i>	10/3/16	12:00	<i>Bl M Pann</i>	10/6/16	10:15	Received on Ice Temp in °C Cooler Custody Sealed Samples Intact

SAMPLER NAME AND SIGNATURE
 PRINT Name of SAMPLER: **James Murphy - PACE Analytical**
 SIGNATURE of SAMPLER: *James Murphy*
 DATE Signed (MM/DD/YYYY): **9/27/16**

CHAIN-OF-CUSTODY / Analytical Request Document

The Chain-of-Custody is a LEGAL DOCUMENT. All relevant fields must be completed accurately.



Page: **34** of **9**

Section A

Required Client Information:

Company: **Hannibal CSD**
 Address: **928 Cayuga St**
 Hannibal, NY 13074
 Email To: **nihenmet@hannibalcsd.org**
 Phone: _____
 Project Name: **Fairly Elementary**
 Project Number: _____

Section B

Required Project Information:

Report To: _____
 Copy To: _____
 Purchase Order No.: _____
 Project Name: **Fairly Elementary**
 Project Number: _____

Section C

Invoice Information:

Attention: _____
 Company Name: _____
 Address: _____
 Pace Quote Reference: _____
 Pace Project Manager: _____
 Pace Profile #: _____

REGULATORY AGENCY

NPDES GROUND WATER DRINKING WATER
 UST RCRA OTHER _____

SITE LOCATION
 GA IL IN MI NC
 OH SC WI OTHER NY

ITEM #	SAMPLE ID (A-Z, 0-9 / .)	Valid Matrix Codes MATRIX DW WT WW WATER WASTE WATER WASTEWATER EXTRACT EXPOSURE OIL WIRE OTHER TERRIE	MATRIX CODE	SAMPLE TYPE	COLLECTED		# OF CONTAINERS	PRESERVATIVES	Filtered (Y/N)	Requested Analyte	Face Project No. Lab ID.
					COMPOSITE START	COMPOSITE END/GRAB					
1	Room 131/BS-RIGHT		DW G	G-GRAB C-COMP	DATE	TIME	1	Unpreserved		X	037
2	Boys Locker Room / LEFT SINK		DW G		9/27	3:10	1				038
3	Boys Locker Room / RIGHT SINK		DW G		9/27	3:12	1				039
4	Boys Gym / OFFICE BATH		DW G		9/27	3:14	1				040
5	Room 132 / BS		DW G		9/27	3:15	1				041
6	MEANS ROOM / ACROSS FROM ROOM 132 / SINK 1		DW G		9/27	3:19	1				042
7	MEANS ROOM / ACROSS FROM ROOM 132 / SINK 2		DW G		9/27	3:20	1				043
8	MEANS ROOM / ACROSS FROM ROOM 132 / SINK 3		DW G		9/27	3:20	1				044
9	MEANS ROOM / ACROSS FROM ROOM 132 / SINK 4		DW G		9/27	3:21	1				045
10	LADIES ROOM / ACROSS FROM 132 / SINK 1		DW G		9/27	3:26	1				046
11	LADIES ROOM / ACROSS FROM ROOM 132 / SINK 2		DW G		9/27	3:26	1				047
12	LADIES ROOM / ACROSS FROM ROOM 132 / SINK 3		DW G		9/27	3:27	1				048

RELINQUISHED BY / AFFILIATION	DATE	TIME	ACCEPTED BY / AFFILIATION	DATE	TIME	TEMP IN °C	Received on Ice	Custody Sealed	Samples Intact
<i>James Murphy</i>	10/3/16	17:00	<i>James Murphy</i>	10/6/16	10:15	5°C	Y/N	Y/N	Y/N
							Y/N	Y/N	Y/N
							Y/N	Y/N	Y/N

SAMPLER NAME AND SIGNATURE
 PRINT Name of SAMPLER: **James Murphy - PACE Analytical**
 SIGNATURE of SAMPLER: *James Murphy*
 DATE Signed (MM/DD/YY): **9/27/16**

CHAIN-OF-CUSTODY / Analytical Request Document

The Chain-of-Custody is a LEGAL DOCUMENT. All relevant fields must be completed accurately.



Section A
Required Client Information:
Company: Hannibal CSD
Address: 928 Cayuga St
Hannibal, NY 13074
Email To: rhenner@hannibalcsd.org
Phone: _____
Requested Due Date/TAT: _____

Section B
Required Project Information:
Report To: _____
Copy To: _____
Purchase Order No.: _____
Project Name: Fairy Elementary
Project Number: _____

Section C
Invoice Information:
Attention: _____
Company Name: _____
Address: _____
Pace Quote Reference: _____
Pace Project Manager: _____
Pace Profile #: _____

REGULATORY AGENCY

NPDES GROUND WATER DRINKING WATER

UST RCRA OTHER _____

SITE GA IL IN MI NC

LOCATION OH SC WI OTHER_NY _____

ITEM #	Valid Matrix Codes MATRIX DRINKING WATER WATER WATER PRODUCT SOLID OIL AIR OTHER TISSUE	CODE DW WW P SL WR AR OT TS	COLLECTED		SAMPLE TYPE G-RAB C-COMP	MATRIX CODE	COLLECTION		# OF CONTAINERS	PRESERVATIVES						Requested Ante	Pace Project No. Lab ID
			COMPOSITE START DATE TIME	COMPOSITE END/GRAB DATE TIME			UNPRESERVED	H2SO4		HNO3	HCl	NaOH	Na2S2O8	Methanol	Na2SO4		
1	LADIES ROOM / ACROSS FROM RM 132/SINK 4	DW	G	9/27 3:27	9/27 3:27	DW	G	1	1							049	
2	ROOM 133/LEFT SINK	DW	G	9/27 3:30	9/27 3:30	DW	G	1	1							050	
3	ROOM 133/RIGHT SINK	DW	G	9/27 3:30	9/27 3:30	DW	G	1	1							051	
4	ROOM 133/BS	DW	G	9/27 3:32	9/27 3:32	DW	G	1	1							052	
5	GIRLS LOCKER ROOM/LEFT SINK	DW	G	9/27 3:34	9/27 3:34	DW	G	1	1							053	
6	GIRLS LOCKER ROOM/RIGHT SINK	DW	G	9/27 3:34	9/27 3:34	DW	G	1	1							054	
7	GIRLS LOCKER BATH	DW	G	9/27 3:31	9/27 3:31	DW	G	1	1							055	
8	Room 134/CS	DW	G	9/27 3:36	9/27 3:36	DW	G	1	1							056	
9	Room 134/BS	DW	G	9/27 3:36	9/27 3:36	DW	G	1	1							057-058	
10	Room 136/CS	DW	G	9/27 3:41	9/27 3:41	DW	G	1	1							059-059	
11	Room 136/BS	DW	G	9/27 3:42	9/27 3:42	DW	G	1	1							060-059	
12	Room 135/CS/BS	DW	G	9/27 3:42	9/27 3:42	DW	G	1	1							060-059	

RELIQUISHED BY / AFFILIATION *JAMES MURPHY PACE* **DATE** 9/27/16

ACCEPTED BY / AFFILIATION *PAUL M. SPURGE* **DATE** 10/6/16

TEMP IN °C 52.2

Temp in °C 52.2

Received on 10/6/16

Custody Sealed Y/N

Samples Intact Y/N

Section A
 Required Client Information:
 Company: Hannibal CSD
 Address: 928 Cayuga St
 Hannibal, NY 13074
 Email To: rhenner@hannibalcsd.org
 Phone: _____
 Fax: _____

Section B
 Required Project Information:
 Report To: _____
 Copy To: _____
 Purchase Order No.: _____
 Project Name: Fairly Elementary
 Project Number: _____

Section C
 Invoice Information:
 Attention: _____
 Company Name: _____
 Address: _____
 Pace Quote Reference: _____
 Pace Project Manager: _____
 Pace Profile #: _____

REGULATORY AGENCY

NPDES GROUND WATER DRINKING WATER

UST RCRA OTHER _____

SITE
 GA IL IN MI NC

LOCATION
 OH SC WI OTHER_NY

#	MATRIX	CODE	SAMPLER TYPE	COLLECTED		# OF CONTAINERS	PRESERVATIVES							Requested Ante	Filtered (Y/N)	Temp in °C	Received on	Custody Sealed	Samples In tact
				COMPOSITE START	COMPOSITE END/GRAB		Unpreserved	H ₂ O ₂	HNO ₃	HCl	NaOH	Na ₂ SO ₄	Methanol						
1	Room 202/CS	DW	G	9:57	3:55	1										X	67.2-073		
2	Room 203/CS	DW	G	3:56	3:56	1											073-074		
3	Room 204/CS	DW	G	3:56	3:56	1											074-075		
4	Room 205/CS	DW	G	3:57	3:57	1											075-076		
5	2nd Flr/Boys Room / SEMI 1	DW	G	3:59	3:59	1											076-077		
6	2nd Flr/Boys Room / SEMI 2	DW	G	4:00	4:00	1											077-078		
7	2nd Flr/Boys Room / SEMI 3	DW	G	4:01	4:01	1											078-079		
8	2nd Flr/Boys Room / SEMI 4	DW	G	4:02	4:02	1											079-080		
9	2nd Flr/Boys Room / SEMI 5	DW	G	4:02	4:02	1											080-081		
10	2nd Flr/Boys Room / SEMI 6	DW	G	4:03	4:03	1											081-082		
11	2nd Flr/ Faculty Bath / LEFC SEMI 6	DW	G	4:06	4:06	1											082-083		
12	2nd Flr/ Faculty Bath / REGUR SEMI 6	DW	G	4:06	4:06	1											083-084		

ADDITIONAL COMMENTS
 JRM/MP PACE 10/3/16 17:00
 JRM/MP PACE 10/6/16 10:15

RELINQUISHED BY / AFFILIATION
 JRM/MP PACE

ACCEPTED BY / AFFILIATION
 JRM/MP PACE

DATE
 10/3/16 17:00

DATE
 10/6/16 10:15

TIME
 17:00

TIME
 10:15

SAMPLER NAME AND SIGNATURE
 James Murphy - PACE Analytical
 PRINT Name of SAMPLER:
 SIGNATURE of SAMPLER: *JRM/MP*

DATE SIGNED (MM/DD/YY): 9/22/16

Section A

Section B

Section C

Required Client Information:
 Company: Hannibal CSD
 Address: 928 Cayuga St
 Hannibal, NY 13074
 Email To: rhenner@hannibalcsd.org
 Phone: []
 Project Name: Fairy Elementary
 Project Number: []

Report To:
 Copy To:
 Purchase Order No.:
 Project Manager:
 Pace Quote Reference:
 Pace Profile #:

REGULATORY AGENCY
 NPDES GROUND WATER DRINKING WATER
 UST RCRA OTHER
SITE LOCATION
 GA IL IN MI NC
 OH SC WI OTHER_NY

ITEM #	Section D Required Client Information	Valid Matrix Codes	MATRIX	CODE	COLLECTED		SAMPLER TYPE	MATRIX CODE	# OF CONTAINERS	PRESERVATIVES						Requested	Pace Project No. Lab I.D.
					COMPOSITE START	COMPOSITE END/GRAB				Unpreserved	H ₂ SO ₄	HNO ₃	HCl	NaOH	Na ₂ S ₂ O ₃		
		DATE		TIME	DATE	TIME	SAMPLER TEMP AT COLLECTION		ACCEPTED BY / AFFILIATION		DATE	TIME	SAMPLE CONDITIONS				
1	ROOM 206 / CS	DW	G	9/27	4:08	9/27	4:08	DW	1							084085	
2	ROOM 207 / CS	DW	G		4:09		4:09	DW	1							065086	
3	2ND FLOOR / MOMENTS FACILITY BATH / SINK	DW	G		4:10		4:10	DW	1							086087	
4	2ND FLOOR / MOMENTS FACILITY BATH / SINK	DW	G		4:11		4:11	DW	1							067088	
5	2ND FLOOR / GIRLS ROOM / SINK 1	DW	G		4:12		4:12	DW	1							088089	
6	2ND FLOOR / GIRLS ROOM / SINK 2	DW	G		4:13		4:13	DW	1							089090	
7	2ND FLOOR / GIRLS ROOM / SINK 3	DW	G		4:14		4:14	DW	1							090091	
8	2ND FLOOR / GIRLS ROOM / SINK 4	DW	G		4:15		4:15	DW	1							091092	
9	2ND FLOOR / GIRLS ROOM / SINK 5	DW	G		4:16		4:16	DW	1							092093	
10	2ND FLOOR / GIRLS ROOM / SINK 6	DW	G		4:17		4:17	DW	1							093094	
11	ROOM 209 / CS	DW	G		4:17		4:17	DW	1							094095	
12	ROOM 208 / CS	DW	G		4:18		4:18	DW	1							095096	

ADDITIONAL COMMENTS
 JMM
 MCE
 9/27/16
 17:00
 PA
 M
 10/6/14
 15:15

RELINQUISHED BY / AFFILIATION
 DATE
 TIME

ACCEPTED BY / AFFILIATION
 DATE
 TIME

SAMPLE CONDITIONS
 Received on ice
 Custody Sealed
 Samples Intact

Temp in °C

SAMPLER NAME AND SIGNATURE
 PRINT Name of SAMPLER: James Murphy - PACE Analytical
 SIGNATURE of SAMPLER: *JMM*
 DATE Signed (MM / DD / YY): 9/27/16

Section A Required Client Information:

Company: Hannibal CSD
 Address: 928 Cayuga St
 Hannibal, NY 13074
 Email To: nhenner@hannibalcsd.org
 Phone: _____
 Project Name: Fatty Elementary
 Project Number: _____

Section B Required Project Information:

Report To: _____
 Copy To: _____
 Purchase Order No.: _____
 Project Manager: _____

Section C Invoice Information:

Attention: _____
 Company Name: _____
 Address: _____
 Pace Quote Reference: _____
 Pace Project Manager: _____
 Pace Profile #: _____

REGULATORY AGENCY

NPDES GROUND WATER DRINKING WATER
 UST RCRA OTHER _____

SITE GA IL IN MI NC
LOCATION OH SC WI OTHER NY _____

ITEM #	Valid Matrix Codes MATRIX DRINKING WATER WASTE WATER WASTEWATER SOLID OIL MISC OTHER TISSE	Section D Required Client Information SAMPLE ID (A-Z, 0-9 / -) Sample IDs MUST BE UNIQUE	CODE DW WT P SL OL M AR OT TS	MATRIX CODE	G-RAB C-COMP	COLLECTED		# OF CONTAINERS	PRESERVATIVES	UNPRESERVED	H ₂ SO ₄	HNO ₃	HCl	NaOH	Na ₂ SO ₄	Methanol	Na ₂ S ₂ O ₈	M ₂ SO ₄	Requested Analyte	Filtered (Y/N)	Pace Project No. Lab I.D.	
						COMPOSITE START DATE TIME	COMPOSITE END/GRAB DATE TIME															
1		Room 210/CS	DW	G		9/27	4:20	1														096097
2		Room 212/CS	DW	G		9/27	4:21	1														097098
3		Room 211/CS	DW	G		9/27	4:21	1														098099
4		Room 213/CS	DW	G		9/27	4:22	1														099101
5		Room 215/CS	DW	G		9/27	4:24	1														102 102
6		Room 214/CS	DW	G		9/27	4:25	1														101 103
7		Room 216/CS	DW	G		9/27	4:25	1														102 104
8			DW	G				1														
9			DW	G				1														
10			DW	G				1														
11			DW	G				1														
12			DW	G				1														


ADDITIONAL COMMENTS	RELINQUISHED BY / AFFILIATION	DATE	TIME	ACCEPTED BY / AFFILIATION	DATE	TIME	SAMPLE CONDITIONS
	James Murphy	10/31/16	17:00	Bel Th	10/6/16	10:15	Received on Ice Temp in C Custody Sealed Cooler Samples Intact

SAMPLER NAME AND SIGNATURE
 PRINT Name of SAMPLER: James Murphy - PACE Analytical
 SIGNATURE of SAMPLER: *James Murphy*
 DATE Signed (MM/DD/YY): 9/27/16

Sample Condition Upon Receipt

Client Name: NY Schools Project #: **WO# : 10365060**

Courier: Fed Ex UPS USPS Client
 Commercial Pace Speedee Other: _____
 Tracking Number: See exception



Custody Seal on Cooler/Box Present? Yes No Seals Intact? Yes No

Optional: Proj. Due Date: _____ Proj. Name: _____

Packing Material: Bubble Wrap Bubble Bags None Other: _____ Temp Blank? Yes No

Thermometer Used: 151401163 151401164 B88A912167504 B88A0143310098
 Type of Ice: Wet Blue None Samples on ice, cooling process has begun

Cooler Temp Read (°C): _____ Cooler Temp Corrected (°C): See exception Biological Tissue Frozen? Yes No N/A
 Temp should be above freezing to 6°C Correction Factor: +0.2 Date and Initials of Person Examining Contents: BSM 10/6/16

USDA Regulated Soil (N/A, water sample)
 Did samples originate in a quarantine zone within the United States: AL, AR, AZ, CA, FL, GA, ID, LA, MS, NC, NM, NY, OK, OR, SC, TN, TX or VA (check maps)? Yes No
 Did samples originate from a foreign source (internationally, including Hawaii and Puerto Rico)? Yes No

If Yes to either question, fill out a Regulated Soil Checklist (F-MN-Q-338) and include with SCUR/COC paperwork.

	COMMENTS:
Chain of Custody Present? <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	1.
Chain of Custody Filled Out? <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	2.
Chain of Custody Relinquished? <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	3.
Sampler Name and/or Signature on COC? <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	4.
Samples Arrived within Hold Time? <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	5.
Short Hold Time Analysis (<72 hr)? <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No <input type="checkbox"/> N/A	6.
Rush Turn Around Time Requested? <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No <input type="checkbox"/> N/A	7.
Sufficient Volume? <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	8.
Correct Containers Used? <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	9.
-Pace Containers Used? <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	
Containers Intact? <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	10.
Filtered Volume Received for Dissolved Tests? <input type="checkbox"/> Yes <input type="checkbox"/> No <input checked="" type="checkbox"/> N/A	11. Note if sediment is visible in the dissolved container
Sample Labels Match COC? <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	12.
-Includes Date/Time/ID/Analysis Matrix: <u>WT</u>	
All containers needing acid/base preservation have been checked? <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	13. <input checked="" type="checkbox"/> HNO ₃ <input type="checkbox"/> H ₂ SO ₄ <input type="checkbox"/> NaOH <input type="checkbox"/> HCl
All containers needing preservation are found to be in compliance with EPA recommendation? (HNO ₃ , H ₂ SO ₄ , HCl<2; NaOH >9 Sulfide, NaOH>12 Cyanide) Exceptions: VOA, Coliform, TOC, Oil and Grease, DRO/8015 (water) DOC <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	Sample # <u>1/1</u> <u>01-103</u>
Headspace in VOA Vials (>6mm)? <input type="checkbox"/> Yes <input type="checkbox"/> No <input checked="" type="checkbox"/> N/A	Initial when completed: _____ Lot # of added preservative: _____
Trip Blank Present? <input type="checkbox"/> Yes <input type="checkbox"/> No <input checked="" type="checkbox"/> N/A	14.
Trip Blank Custody Seals Present? <input type="checkbox"/> Yes <input type="checkbox"/> No <input checked="" type="checkbox"/> N/A	15.
Pace Trip Blank Lot # (if purchased): _____	

CLIENT NOTIFICATION/RESOLUTION

Field Data Required? Yes No

Person Contacted: _____ Date/Time: _____

Comments/Resolution: _____

Project Manager Review: _____ Date: 10/07/16

Note: Whenever there is a discrepancy affecting North Carolina compliance samples, a copy of this form will be sent to the North Carolina DEHNR Certification Office (i.e. out of hold, incorrect preservative, out of temp, incorrect containers).

